

# MINOTROL CRP

## Hematology Reference Control

9755 Research Drive  
 Irvine, California 92618  
 Telephone: (949) 453-0500  
 (888) 903-5001  
 Fax: (949) 753-0533

Lot No.: **MC410**  
 Exp. Date: **2019-04-05** (Y/M/D)

	Parameter/Units	Low Abnormal Lot # <b>MC410-1</b>			Normal Lot # <b>MC410-2</b>			High Abnormal Lot # <b>MC410-3</b>		
		Assay	Range	Limits	Assay	Range	Limits	Assay	Range	Limits
MICROS CRP 200	WBC K/ $\mu$ L	<b>2.0</b>	0.4	1.6 - 2.4	<b>7.6</b>	1.0	6.6 - 8.6	<b>19.9</b>	2.2	17.7 - 22.1
	RBC M/ $\mu$ L	<b>2.37</b>	0.15	2.22 - 2.52	<b>4.75</b>	0.18	4.57 - 4.93	<b>5.73</b>	0.25	5.48 - 5.98
	HGB g/dL	<b>6.1</b>	0.5	5.6 - 6.6	<b>13.5</b>	0.6	12.9 - 14.1	<b>18.0</b>	0.7	17.3 - 18.7
	HCT %	<b>16.6</b>	2.0	14.6 - 18.6	<b>38.0</b>	2.5	35.5 - 40.5	<b>51.0</b>	3.0	48.0 - 54.0
	MCV fL	<b>70</b>	4	66 - 74	<b>80</b>	4	76 - 84	<b>89</b>	4	85 - 93
	MCH pg	<b>25.7</b>	2.4	23.3 - 28.1	<b>28.4</b>	2.4	26.0 - 30.8	<b>31.4</b>	2.8	28.6 - 34.2
	MCHC g/dL	<b>36.8</b>	3.0	33.8 - 39.8	<b>35.5</b>	3.0	32.5 - 38.5	<b>35.3</b>	3.0	32.3 - 38.3
	RDW %	<b>13.0</b>	4.0	9.0 - 17.0	<b>13.5</b>	4.0	9.5 - 17.5	<b>13.0</b>	4.0	9.0 - 17.0
	PLT K/ $\mu$ L	<b>72</b>	20	52 - 92	<b>265</b>	40	225 - 305	<b>515</b>	60	455 - 575
	MPV fL	<b>8.4</b>	2.0	6.4 - 10.4	<b>8.4</b>	2.0	6.4 - 10.4	<b>8.0</b>	2.0	6.0 - 10.0
	Lym %	<b>65.0</b>	8.0	57.0 - 73.0	<b>34.0</b>	6.0	28.0 - 40.0	<b>15.0</b>	6.0	9.0 - 21.0
	Lym K/ $\mu$ L	<b>1.3</b>	0.2	1.1 - 1.5	<b>2.6</b>	0.5	2.1 - 3.1	<b>3.0</b>	1.2	1.8 - 4.2
	Mon %	<b>9.5</b>	7.0	2.5 - 16.5	<b>8.0</b>	5.0	3.0 - 13.0	<b>6.0</b>	5.0	1.0 - 11.0
	Mon K/ $\mu$ L	<b>0.2</b>	0.2	0.0 - 0.4	<b>0.6</b>	0.4	0.2 - 1.0	<b>1.2</b>	1.0	0.2 - 2.2
	Gran %	<b>25.5</b>	7.0	18.5 - 32.5	<b>58.0</b>	6.0	52.0 - 64.0	<b>79.0</b>	7.0	72.0 - 86.0
	Gran K/ $\mu$ L	<b>0.5</b>	0.2	0.3 - 0.7	<b>4.4</b>	0.5	3.9 - 4.9	<b>15.7</b>	1.4	14.3 - 17.1
	CRP mg/dL	<b>0.52</b>	0.30	0.22 - 0.82	<b>1.95</b>	0.70	1.25 - 2.65	<b>3.57</b>	1.00	2.57 - 4.57

For technical assistance, call the Customer Support Center at 1-888-903-5001.