

# ABX Minocal



**LOT**

**CX 477**  
Rev 1

**CAL**

(Exp.) **2023-01-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.2</b>	<b>9.1</b>	<b>9.2</b>	<b>9.0</b>	<b>9.2</b>	<b>8.9</b>	<b>9.2</b>	<b>9.1</b>	<b>9.0</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.58</b>	<b>4.54</b>	<b>4.54</b>	<b>4.56</b>	<b>4.58</b>	<b>4.51</b>	<b>4.58</b>	<b>4.58</b>	<b>4.56</b>	± 0.06
HB HGB	g/dl	<b>13.9</b>	<b>13.8</b>	<b>13.8</b>	<b>14.0</b>	<b>14.0</b>	<b>13.9</b>	<b>14.0</b>	<b>14.0</b>	<b>14.0</b>	± 0.2
	g/l	<b>139</b>	<b>138</b>	<b>138</b>	<b>140</b>	<b>140</b>	<b>139</b>	<b>140</b>	<b>140</b>	<b>140</b>	± 2
	mmol/l	<b>8.63</b>	<b>8.57</b>	<b>8.57</b>	<b>8.69</b>	<b>8.69</b>	<b>8.63</b>	<b>8.69</b>	<b>8.69</b>	<b>8.69</b>	± 0.12
HT HCT	%	<b>39.4</b>	<b>39.0</b>	<b>38.4</b>	<b>38.3</b>	<b>38.2</b>	<b>37.9</b>	<b>38.2</b>	<b>38.7</b>	<b>38.3</b>	± 1.0
	l/l	<b>0.394</b>	<b>0.390</b>	<b>0.384</b>	<b>0.383</b>	<b>0.382</b>	<b>0.379</b>	<b>0.382</b>	<b>0.387</b>	<b>0.383</b>	± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>260</b>	<b>254</b>	<b>246</b>	<b>245</b>	<b>257</b>	<b>250</b>	<b>257</b>	<b>249</b>	<b>245</b>	± 10
VMP MPV	µm <sup>3</sup> ; fl	<b>8.9</b>	<b>8.9</b>	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.48 BACK / VERSO 1300090308