



LOT

MX 439 L

CONTROL L



2023-04-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	1.9			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.35			2.30					± 0.17
HB HGB	g/dl	6.1	6.0			5.8					± 0.5
	g/l	61	60			58					± 5
	mmol/l	3.79	3.71			3.58					± 0.31
HT HCT	%	16.8	16.9			16.8					± 2.0
	l/l	0.168	0.169			0.168					± 0.020
VGM MCV	µm ³ ; fl	72	72			73					± 4
TGMH MCH	pg	26.2	25.4			25.1					± 2.0
	fmol	1.63	1.58			1.56					± 0.12
CCMH MCHC	g/dl	36.4	35.3			34.3					± 3.0
	g/l	364	353			343					± 30
	mmol/l	22.58	21.94			21.33					± 1.86
IDR RDW	%	15.5	14.0			15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	76	79			76					± 22
VPM MPV	µm ³ ; fl	8.9	8.5			8.4					± 2.0
LYMPHO	#	1.31	1.25			1.32					± 0.40
	%	65.5	65.7			66.0					± 8.0
MONO	#	0.16	0.17			0.15					± 0.20
	%	8.0	9.0			7.5					± 6.0
GRANULO	#	0.53	0.48			0.53					± 0.40
	%	26.5	25.5			26.5					± 7.0

LOT		MX 439 N		CONTROL N						(Exp.) 2023-04-05 (YYYY-MM-DD)	
Rev 1		Milysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.3			7.4					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.50	4.52			4.45					± 0.26
HB HGB	g/dl	13.5	13.5			13.1					± 0.7
	g/l	135	135			131					± 7
	mmol/l	8.38	8.38			8.15					± 0.43
HT HCT	%	36.9	37.1			36.9					± 2.7
	l/l	0.369	0.371			0.369					± 0.027
VGM MCV	µm ³ ; fl	82	82			83					± 4
TGMH MCH	pg	30.0	29.9			29.5					± 2.0
	fmol	1.86	1.85			1.83					± 0.12
CCMH MCHC	g/dl	36.6	36.4			35.6					± 3.0
	g/l	366	364			356					± 30
	mmol/l	22.72	22.62			22.08					± 1.86
IDR RDW	%	15.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	260	272			268					± 40
VPM MPV	µm ³ ; fl	8.5	8.2			7.9					± 2.0
LYMPHO	#	2.59	2.59			2.55					± 0.40
	%	34.5	35.5			34.5					± 5.0
MONO	#	0.53	0.55			0.52					± 0.30
	%	7.0	7.6			7.0					± 4.0
GRANULO	#	4.39	4.15			4.33					± 0.60
	%	58.5	56.8			58.5					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143



LOT

MX 439 H

CONTROL

H

(Exp.)

2023-04-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	19.7	19.6			19.7					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.49	5.50			5.50					± 0.32
HB HGB	g/dl	18.2	17.6			17.5					± 0.9
	g/l	182	176			175					± 9
	mmol/l	11.30	10.93			10.88					± 0.56
HT HCT	%	49.4	49.0			50.1					± 3.4
	l/l	0.494	0.490			0.501					± 0.034
VGM MCV	µm ³ ; fl	90	89			91					± 4
TGMH MCH	pg	33.2	32.0			31.9					± 2.0
	fmol	2.06	1.99			1.98					± 0.12
CCMH MCHC	g/dl	36.8	36.0			35.0					± 3.0
	g/l	368	360			350					± 30
	mmol/l	22.87	22.33			21.74					± 1.86
IDR RDW	%	14.0	13.0			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	485	525			516					± 65
VPM MPV	µm ³ ; fl	8.4	8.2			7.8					± 2.0
LYMPHO	#	3.45	3.43			3.45					± 1.00
	%	17.5	17.5			17.5					± 5.0
MONO	#	1.08	1.27			1.08					± 0.80
	%	5.5	6.5			5.5					± 4.0
GRANULO	#	15.17	14.89			15.17					± 1.40
	%	77.0	76.0			77.0					± 7.0