

# ABX Minotrol CRP



**LOT**

**MC 424**

**CONTROL**



**(Exp.) 2022-10-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														
				CONTROL				1	CONTROL				2	CONTROL				3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE
				EMI CRP	LC-767G	MS CRP	EMI CRP		LC-767G	MS CRP	EMI CRP	LC-767G		MS CRP				
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>	<b>2.1</b>	± 0.4	<b>7.6</b>	<b>7.7</b>	<b>7.7</b>	± 1.0	<b>20.0</b>	<b>20.4</b>	<b>20.6</b>	± 2.2				
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.26</b>	<b>2.28</b>	<b>2.22</b>	± 0.15	<b>4.46</b>	<b>4.48</b>	<b>4.47</b>	± 0.18	<b>5.50</b>	<b>5.53</b>	<b>5.59</b>	± 0.25				
HB	HGB	g/dl	<b>6.0</b>	<b>6.0</b>	<b>5.7</b>	± 0.4	<b>13.6</b>	<b>13.5</b>	<b>13.5</b>	± 0.5	<b>17.6</b>	<b>17.9</b>	<b>18.1</b>	± 0.7				
		g/l	<b>60</b>	<b>60</b>	<b>57</b>	± 4	<b>136</b>	<b>135</b>	<b>135</b>	± 5	<b>176</b>	<b>179</b>	<b>181</b>	± 7				
		mmol/l	<b>3.73</b>	<b>3.73</b>	<b>3.54</b>	± 0.25	<b>8.45</b>	<b>8.38</b>	<b>8.38</b>	± 0.31	<b>10.93</b>	<b>11.12</b>	<b>11.24</b>	± 0.44				
HT	HCT	%	<b>16.3</b>	<b>16.4</b>	<b>15.8</b>	± 2.0	<b>37.0</b>	<b>37.6</b>	<b>36.7</b>	± 2.5	<b>49.0</b>	<b>49.8</b>	<b>49.2</b>	± 3.0				
		l/l	<b>0.163</b>	<b>0.164</b>	<b>0.158</b>	± 0.020	<b>0.370</b>	<b>0.376</b>	<b>0.367</b>	± 0.025	<b>0.490</b>	<b>0.498</b>	<b>0.492</b>	± 0.030				
VGM	MCV	µm <sup>3</sup> ; fl	<b>72</b>	<b>72</b>	<b>71.0</b>	± 4	<b>83</b>	<b>84</b>	<b>82.0</b>	± 4	<b>89</b>	<b>90</b>	<b>88.0</b>	± 4				
TGMH	MCH	pg	<b>26.5</b>	<b>26.3</b>	<b>25.7</b>	± 2.4	<b>30.5</b>	<b>30.1</b>	<b>30.2</b>	± 2.4	<b>32.0</b>	<b>32.4</b>	<b>32.4</b>	± 2.8				
		fmol	<b>1.65</b>	<b>1.63</b>	<b>1.59</b>	± 0.15	<b>1.89</b>	<b>1.87</b>	<b>1.88</b>	± 0.15	<b>1.99</b>	<b>2.01</b>	<b>2.01</b>	± 0.18				
CCMH	MCHC	g/dl	<b>36.9</b>	<b>36.5</b>	<b>36.2</b>	± 3.0	<b>36.7</b>	<b>35.9</b>	<b>36.8</b>	± 3.0	<b>36.0</b>	<b>36.0</b>	<b>36.8</b>	± 3.0				
		g/l	<b>369</b>	<b>365</b>	<b>362</b>	± 30	<b>367</b>	<b>359</b>	<b>368</b>	± 30	<b>360</b>	<b>360</b>	<b>368</b>	± 30				
		mmol/l	<b>22.90</b>	<b>22.70</b>	<b>22.46</b>	± 1.86	<b>22.81</b>	<b>22.28</b>	<b>22.87</b>	± 1.86	<b>22.33</b>	<b>22.33</b>	<b>22.85</b>	± 1.86				
IDR	RDW	%	<b>13.5</b>	<b>13.0</b>	<b>15.0</b>	± 3.0	<b>12.5</b>	<b>12.5</b>	<b>12.5</b>	± 3.0	<b>12.0</b>	<b>12.0</b>	<b>12.0</b>	± 3.0				
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>77</b>	<b>79</b>	<b>68</b>	± 20	<b>271</b>	<b>263</b>	<b>258</b>	± 40	<b>490</b>	<b>492</b>	<b>500</b>	± 60				
VPM	MPV	µm <sup>3</sup> ; fl	<b>7.7</b>	<b>8.4</b>	<b>9.0</b>	± 2.0	<b>8.3</b>	<b>8.3</b>	<b>9.2</b>	± 2.0	<b>8.1</b>	<b>8.3</b>	<b>8.7</b>	± 2.0				
LYMPHO		#	<b>1.3</b>	<b>1.3</b>	N/A	± 0.2	<b>2.7</b>	<b>2.7</b>	N/A	± 0.5	<b>3.4</b>	<b>3.3</b>	N/A	± 1.1				
		%	<b>62.0</b>	<b>61.5</b>	N/A	± 8.0	<b>35.0</b>	<b>34.5</b>	N/A	± 6.0	<b>17.0</b>	<b>16.0</b>	N/A	± 5.0				
MONO		#	<b>0.2</b>	<b>0.2</b>	N/A	± 0.2	<b>0.6</b>	<b>0.6</b>	N/A	± 0.4	<b>1.3</b>	<b>1.2</b>	N/A	± 1.0				
		%	<b>9.0</b>	<b>9.0</b>	N/A	± 7.0	<b>8.0</b>	<b>8.0</b>	N/A	± 5.0	<b>6.5</b>	<b>6.0</b>	N/A	± 5.0				
GRANULO		#	<b>0.6</b>	<b>0.6</b>	N/A	± 0.2	<b>4.3</b>	<b>4.4</b>	N/A	± 0.5	<b>15.3</b>	<b>15.9</b>	N/A	± 1.5				
		%	<b>29.0</b>	<b>29.5</b>	N/A	± 7.0	<b>57.0</b>	<b>57.5</b>	N/A	± 6.0	<b>76.5</b>	<b>78.0</b>	N/A	± 7.0				
CRP		mg/dl	<b>0.46</b>	<b>0.51</b>	<b>0.49</b>	± 0.40	<b>1.88</b>	<b>2.07</b>	<b>1.96</b>	± 0.80	<b>3.34</b>	<b>3.66</b>	<b>3.57</b>	± 1.50				
		mg/l	<b>4.6</b>	<b>5.1</b>	<b>4.9</b>	± 4.0	<b>18.8</b>	<b>20.7</b>	<b>19.6</b>	± 8.0	<b>33.4</b>	<b>36.6</b>	<b>35.7</b>	± 15.0				

Ref: TEMP-1070 Rev.42 BACK/VERSO 1300092170