

ABX Minocal



LOT

CX 479

CAL



(Exp.) 2023-03-05
(YYYY - MM - DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.0	9.0	9.2	8.9	9.1	8.8	9.1	9.0	8.9	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.55	4.49	4.48	4.52	4.55	4.51	4.55	4.57	4.52	± 0.06
HB HGB	g/dl	13.4	13.3	13.3	13.4	13.4	13.4	13.4	13.4	13.4	± 0.2
	g/l	134	133	133	134	134	134	134	134	134	± 2
	mmol/l	8.32	8.26	8.26	8.32	8.32	8.32	8.32	8.32	8.32	± 0.12
HT HCT	%	38.0	37.5	37.0	37.1	36.9	36.8	36.9	37.0	37.1	± 1.0
	l/l	0.380	0.375	0.370	0.371	0.369	0.368	0.369	0.370	0.371	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	258	253	242	242	252	245	252	250	242	± 10
VMP MPV	µm ³ ; fl	9.3	9.3	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.48 BACK / VERSO 1300090308