

LOT MX 440 L
Rev 1

CONTROL L

 (Exp.)

2023-06-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 BP - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.0	2.1			2.0					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.37	2.35			2.37					± 0.17
HB HGB	g/dl	5.9	6.0			5.7					± 0.5
	g/l	59	60			57					± 5
	mmol/l	3.66	3.71			3.52					± 0.31
HT HCT	%	16.6	16.5			17.1					± 2.0
	l/l	0.166	0.165			0.171					± 0.020
VGM MCV	µm ³ ; fl	70	70			72					± 4
TGMH MCH	pg	24.9	25.4			23.9					± 2.0
	fmol	1.55	1.58			1.49					± 0.12
CCMH MCHC	g/dl	35.6	36.3			33.2					± 3.0
	g/l	356	363			332					± 30
	mmol/l	22.08	22.57			20.65					± 1.86
IDR RDW	%	15.5	14.0			15.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	79	84			81					± 22
VPM MPV	µm ³ ; fl	9.0	8.4			8.2					± 2.0
LYMPHO	#	1.36	1.42			1.36					± 0.40
	%	68.0	67.7			68.0					± 8.0
MONO	#	0.13	0.19			0.14					± 0.20
	%	6.5	9.0			7.0					± 6.0
GRANULO	#	0.51	0.49			0.50					± 0.40
	%	25.5	23.5			25.0					± 7.0

LOT		MX 440	N	CONTROL		N					(Exp.)	2023-06-05 (YYYY-MM-DD)
Rev 1												
		Minilysebio										
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE	
		ES60	CRP 200			60 8P - 60 18P						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.6			7.5					± 0.8	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.62	4.58			4.59					± 0.26	
HB HGB	g/dl	13.7	13.8			13.4					± 0.7	
	g/l	137	138			134					± 7	
	mmol/l	8.51	8.57			8.34					± 0.43	
HT HCT	%	37.4	37.6			38.6					± 2.7	
	l/l	0.374	0.376			0.386					± 0.027	
VGM MCV	µm ³ ; fl	81	82			84					± 4	
TGMH MCH	pg	29.7	30.1			29.3					± 2.0	
	fmol	1.84	1.87			1.82					± 0.12	
CCMH MCHC	g/dl	36.6	36.7			34.8					± 3.0	
	g/l	366	367			348					± 30	
	mmol/l	22.73	22.82			21.62					± 1.86	
IDR RDW	%	14.5	13.5			14.5					± 3.0	
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	257	272			267					± 40	
VPM MPV	µm ³ ; fl	8.5	7.9			7.7					± 2.0	
LYMPHO	#	2.77	2.82			2.63					± 0.40	
	%	36.0	37.1			35.0					± 5.0	
MONO	#	0.46	0.47			0.45					± 0.30	
	%	6.0	6.2			6.0					± 4.0	
GRANULO	#	4.47	4.32			4.43					± 0.60	
	%	58.0	56.8			59.0					± 6.0	

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

LOT		MX 440 H	CONTROL H				(Exp.)		2023-06-05 (YYYY-MM-DD)		
Rev 1		Minilysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.3	20.1			19.8					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.69	5.67			5.76					± 0.32
HB HGB	g/dl	18.5	18.2			18.0					± 0.9
	g/l	185	182			180					± 9
	mmol/l	11.49	11.30			11.19					± 0.56
HT HCT	%	50.6	50.5			52.4					± 3.4
	l/l	0.506	0.505			0.524					± 0.034
VGM MCV	µm ³ ; fl	89	89			91					± 4
TGMH MCH	pg	32.5	32.1			31.3					± 2.0
	fmol	2.02	1.99			1.94					± 0.12
CCMH MCHC	g/dl	36.5	36.1			34.4					± 3.0
	g/l	365	361			344					± 30
	mmol/l	22.69	22.40			21.35					± 1.86
IDR RDW	%	14.0	14.0			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	514	537			542					± 65
VPM MPV	µm ³ ; fl	8.5	7.8			7.5					± 2.0
LYMPHO	#	3.65	3.62			3.47					± 1.00
	%	18.0	18.0			17.5					± 5.0
MONO	#	1.02	1.41			0.89					± 0.80
	%	5.0	7.0			4.5					± 4.0
GRANULO	#	15.63	15.07			15.44					± 1.40
	%	77.0	75.0			78.0					± 7.0