

# ABX Minocal



**LOT**

**CX 482**

**CAL**



**(Exp.) 2023-06-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.2</b>	<b>9.4</b>	<b>9.3</b>	<b>9.1</b>	<b>9.4</b>	<b>9.2</b>	<b>9.4</b>	<b>9.2</b>	<b>9.1</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.39</b>	<b>4.38</b>	<b>4.40</b>	<b>4.35</b>	<b>4.41</b>	<b>4.34</b>	<b>4.41</b>	<b>4.38</b>	<b>4.35</b>	± 0.06
HB HGB	g/dl	<b>13.1</b>	<b>13.0</b>	<b>13.1</b>	<b>13.2</b>	<b>13.1</b>	<b>13.1</b>	<b>13.1</b>	<b>13.2</b>	<b>13.2</b>	± 0.2
	g/l	<b>131</b>	<b>130</b>	<b>131</b>	<b>132</b>	<b>131</b>	<b>131</b>	<b>131</b>	<b>132</b>	<b>132</b>	± 2
	mmol/l	<b>8.14</b>	<b>8.07</b>	<b>8.14</b>	<b>8.20</b>	<b>8.14</b>	<b>8.14</b>	<b>8.14</b>	<b>8.20</b>	<b>8.20</b>	± 0.12
HT HCT	%	<b>36.9</b>	<b>36.8</b>	<b>35.9</b>	<b>36.1</b>	<b>35.9</b>	<b>35.6</b>	<b>35.9</b>	<b>35.9</b>	<b>36.1</b>	± 1.0
	l/l	<b>0.369</b>	<b>0.368</b>	<b>0.359</b>	<b>0.361</b>	<b>0.359</b>	<b>0.356</b>	<b>0.359</b>	<b>0.359</b>	<b>0.361</b>	± 0.010
PLA PLT	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>253</b>	<b>247</b>	<b>238</b>	<b>243</b>	<b>253</b>	<b>249</b>	<b>253</b>	<b>249</b>	<b>243</b>	± 10
VMP MPV	µm <sup>3</sup> ; fl	<b>8.5</b>	<b>8.5</b>	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.48 BACK / VERSO 1300090308