

ABX Minocal



LOT

CX 484

CAL



(Exp.) 2023-08-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------------|--------------|--------------|----------------------|--------------|---------------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | LC-767G | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | DX120 - DF120 | 60 60 C+ ES60 | XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.9 | 9.0 | 9.1 | 8.9 | 9.2 | 9.0 | 9.2 | 9.1 | 8.9 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.48 | 4.47 | 4.48 | 4.53 | 4.56 | 4.49 | 4.56 | 4.54 | 4.53 | ± 0.06 |
| HB HGB | g/dl | 13.4 | 13.4 | 13.5 | 13.6 | 13.6 | 13.5 | 13.6 | 13.6 | 13.6 | ± 0.2 |
| | g/l | 134 | 134 | 135 | 136 | 136 | 135 | 136 | 136 | 136 | ± 2 |
| | mmol/l | 8.32 | 8.32 | 8.38 | 8.45 | 8.45 | 8.38 | 8.45 | 8.45 | 8.45 | ± 0.12 |
| HT HCT | % | 38.1 | 38.0 | 37.0 | 37.6 | 37.4 | 36.8 | 37.4 | 37.5 | 37.6 | ± 1.0 |
| | l/l | 0.381 | 0.380 | 0.370 | 0.376 | 0.374 | 0.368 | 0.374 | 0.375 | 0.376 | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 249 | 249 | 234 | 243 | 253 | 245 | 253 | 249 | 243 | ± 10 |
| VMP MPV | µm ³ ; fl | 8.8 | 8.8 | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.49 BACK / VERSO 1300090308