

**LOT** MX 443 L  
Rev 1

**CONTROL** L

 (Exp.)

**2023-12-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.0	2.0			2.1					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.34	2.32			2.31					± 0.17
HB HGB	g/dl	6.0	6.0			5.7					± 0.5
	g/l	60	60			57					± 5
	mmol/l	3.73	3.71			3.52					± 0.31
HT HCT	%	16.6	16.5			16.4					± 2.0
	l/l	0.166	0.165			0.164					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71	71			71					± 4
TGMH MCH	pg	25.6	25.8			24.6					± 2.0
	fmol	1.59	1.60			1.53					± 0.12
CCMH MCHC	g/dl	36.1	36.3			34.6					± 3.0
	g/l	361	363			346					± 30
	mmol/l	22.43	22.54			21.48					± 1.86
IDR RDW	%	15.0	14.0			15.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	74	75			76					± 22
VPM MPV	µm <sup>3</sup> ; fl	8.3	8.0			7.5					± 2.0
LYMPHO	#	1.28	1.29			1.35					± 0.40
	%	67.0	65.7			67.0					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	7.0	9.0			7.0					± 6.0
GRANULO	#	0.52	0.51			0.55					± 0.40
	%	26.0	25.5			26.0					± 7.0

**LOT** MX 443 N  
 Rev 1

**CONTROL** N

(Exp.)

**2023-12-05**  
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.3			7.3					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.57	4.48			4.52					± 0.26
HB HGB	g/dl	13.5	13.4			13.2					± 0.7
	g/l	135	134			132					± 7
	mmol/l	8.38	8.32			8.22					± 0.43
HT HCT	%	37.5	36.7			37.5					± 2.7
	l/l	0.375	0.367			0.375					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	82	82			83					± 4
TGMH MCH	pg	29.5	29.9			29.3					± 2.0
	fmol	1.83	1.86			1.82					± 0.12
CCMH MCHC	g/dl	36.0	36.5			35.3					± 3.0
	g/l	360	365			353					± 30
	mmol/l	22.37	22.65			21.90					± 1.86
IDR RDW	%	14.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	243	251			246					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.2	7.5			7.5					± 2.0
LYMPHO	#	2.49	2.55			2.48					± 0.40
	%	34.5	35.0			34.0					± 5.0
MONO	#	0.45	0.49			0.44					± 0.30
	%	6.0	6.7			6.0					± 4.0
GRANULO	#	4.46	4.26			4.38					± 0.60
	%	59.5	58.3			60.0					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143



<b>LOT</b>		<b>MX 443 H</b>	<b>CONTROL H</b>				(Exp.)				<b>2023-12-05</b> (YYYY-MM-DD)
Rev 1		<b>Milysebio</b>									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.2</b>	<b>20.2</b>			<b>19.6</b>					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.75</b>	<b>5.66</b>			<b>5.68</b>					± 0.32
HB HGB	g/dl	<b>18.5</b>	<b>18.0</b>			<b>17.8</b>					± 0.9
	g/l	<b>185</b>	<b>180</b>			<b>178</b>					± 9
	mmol/l	<b>11.49</b>	<b>11.18</b>			<b>11.07</b>					± 0.56
HT HCT	%	<b>51.2</b>	<b>49.8</b>			<b>50.6</b>					± 3.4
	l/l	<b>0.512</b>	<b>0.498</b>			<b>0.506</b>					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>88</b>			<b>89</b>					± 4
TGMH MCH	pg	<b>32.2</b>	<b>31.8</b>			<b>31.4</b>					± 2.0
	fmol	<b>2.00</b>	<b>1.97</b>			<b>1.95</b>					± 0.12
CCMH MCHC	g/dl	<b>36.2</b>	<b>36.1</b>			<b>35.3</b>					± 3.0
	g/l	<b>362</b>	<b>361</b>			<b>353</b>					± 30
	mmol/l	<b>22.45</b>	<b>22.44</b>			<b>21.89</b>					± 1.86
IDR RDW	%	<b>13.5</b>	<b>13.0</b>			<b>14.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>469</b>	<b>481</b>			<b>488</b>					± 65
VPM MPV	µm <sup>3</sup> ; fl	<b>8.0</b>	<b>7.5</b>			<b>7.0</b>					± 2.0
LYMPHO	#	<b>3.64</b>	<b>3.64</b>			<b>3.53</b>					± 1.00
	%	<b>18.0</b>	<b>18.0</b>			<b>18.0</b>					± 5.0
MONO	#	<b>1.01</b>	<b>1.31</b>			<b>0.88</b>					± 0.80
	%	<b>5.0</b>	<b>6.5</b>			<b>4.5</b>					± 4.0
GRANULO	#	<b>15.55</b>	<b>15.24</b>			<b>15.19</b>					± 1.40
	%	<b>77.0</b>	<b>75.5</b>			<b>77.5</b>					± 7.0