

LOT		MX 444 L		CONTROL L								(Exp.) 2024-02-05 (YYYY-MM-DD)				
		Rev 1		Minilysebio												
PARAMETRES PARAMETERS	UNITES UNITS	MICROS		MICROS		MICROS		MICROS		MICROS		ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200					60 8P - 60 18P								
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	1.9	1.9					2.0								± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.23	2.23					2.24								± 0.17
HB HGB	g/dl	5.7	5.7					5.5								± 0.5
	g/l	57	57					55								± 5
	mmol/l	3.54	3.53					3.41								± 0.31
HT HCT	%	15.8	15.6					15.9								± 2.0
	l/l	0.158	0.156					0.159								± 0.020
VGM MCV	µm ³ ; fl	71	70					71								± 4
TGMH MCH	pg	25.6	25.5					24.5								± 2.0
	fmol	1.59	1.58					1.52								± 0.12
CCMH MCHC	g/dl	36.0	36.4					34.5								± 3.0
	g/l	360	364					345								± 30
	mmol/l	22.36	22.61					21.42								± 1.86
IDR RDW	%	15.0	13.5					15.5								± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	74					77								± 22
VPM MPV	µm ³ ; fl	8.7	8.5					7.9								± 2.0
LYMPHO	#	1.20	1.23					1.29								± 0.40
	%	66.0	66.2					67.0								± 8.0
MONO	#	0.20	0.20					0.20								± 0.20
	%	7.5	9.0					7.5								± 6.0
GRANULO	#	0.50	0.47					0.51								± 0.40
	%	26.5	25.0					25.5								± 7.0

LOT		MX 444 N	CONTROL N				(Exp.)		2024-02-05 (YYYY-MM-DD)		
Rev 1		Minilysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.4	7.2			7.2					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.55	4.51			4.55					± 0.26
HB HGB	g/dl	13.5	13.3			13.2					± 0.7
	g/l	135	133			132					± 7
	mmol/l	8.38	8.26			8.22					± 0.43
HT HCT	%	36.9	36.5			36.9					± 2.7
	l/l	0.369	0.365			0.369					± 0.027
VGM MCV	µm ³ ; fl	81	81			81					± 4
TGMH MCH	pg	29.7	29.5			29.1					± 2.0
	fmol	1.84	1.83			1.81					± 0.12
CCMH MCHC	g/dl	36.6	36.4			35.9					± 3.0
	g/l	366	364			359					± 30
	mmol/l	22.75	22.61			22.29					± 1.86
IDR RDW	%	14.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	243	249			256					± 40
VPM MPV	µm ³ ; fl	8.1	7.9			7.3					± 2.0
LYMPHO	#	2.63	2.56			2.56					± 0.40
	%	35.5	35.5			35.5					± 5.0
MONO	#	0.48	0.51			0.43					± 0.30
	%	6.5	7.1			6.0					± 4.0
GRANULO	#	4.29	4.13			4.21					± 0.60
	%	58.0	57.3			58.5					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

LOT		MX 444 H	CONTROL H				(Exp.)				2024-02-05 (YYYY-MM-DD)
Rev 1		Minilysebio									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	19.9	19.4			19.5					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.66	5.62			5.71					± 0.32
HB HGB	g/dl	18.3	17.8			17.9					± 0.9
	g/l	183	178			179					± 9
	mmol/l	11.36	11.05			11.13					± 0.56
HT HCT	%	50.4	49.5			50.2					± 3.4
	l/l	0.504	0.495			0.502					± 0.034
VGM MCV	µm ³ ; fl	89	88			88					± 4
TGMH MCH	pg	32.3	31.7			31.4					± 2.0
	fmol	2.01	1.97			1.95					± 0.12
CCMH MCHC	g/dl	36.3	36.0			35.7					± 3.0
	g/l	363	360			357					± 30
	mmol/l	22.56	22.35			22.15					± 1.86
IDR RDW	%	14.0	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	503	516			518					± 65
VPM MPV	µm ³ ; fl	8.1	7.8			7.4					± 2.0
LYMPHO	#	3.38	3.29			3.22					± 1.00
	%	17.0	17.0			16.5					± 5.0
MONO	#	0.90	1.36			0.98					± 0.80
	%	4.5	7.0			5.0					± 4.0
GRANULO	#	15.62	14.73			15.31					± 1.40
	%	78.5	76.0			78.5					± 7.0