

# ABX Minotrol CRP



**LOT**

**MC 430**

**CONTROL**



**(Exp.) 2024-04-05**  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio														TOLERANCES TOLERANCE		
				CONTROL				1	CONTROL				2	CONTROL						3
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				TOLERANCES TOLERANCE
				EMI CRP	LC-767G	MS CRP	EMI CRP		LC-767G	MS CRP	EMI CRP	LC-767G		MS CRP						
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>	<b>2.2</b>		± 0.4	<b>7.9</b>	<b>7.6</b>	<b>8.2</b>		± 1.0	<b>20.7</b>	<b>20.4</b>	<b>21.5</b>		± 2.2			
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.25</b>	<b>2.24</b>	<b>2.23</b>		± 0.15	<b>4.50</b>	<b>4.44</b>	<b>4.54</b>		± 0.18	<b>5.48</b>	<b>5.50</b>	<b>5.63</b>		± 0.25			
HB	HGB	g/dl	<b>5.9</b>	<b>5.9</b>	<b>5.7</b>		± 0.4	<b>13.3</b>	<b>13.3</b>	<b>13.5</b>		± 0.5	<b>17.6</b>	<b>17.8</b>	<b>18.1</b>		± 0.7			
		g/l	<b>59</b>	<b>59</b>	<b>57</b>		± 4	<b>133</b>	<b>133</b>	<b>135</b>		± 5	<b>176</b>	<b>178</b>	<b>181</b>		± 7			
		mmol/l	<b>3.66</b>	<b>3.66</b>	<b>3.54</b>		± 0.25	<b>8.26</b>	<b>8.26</b>	<b>8.38</b>		± 0.31	<b>10.93</b>	<b>11.05</b>	<b>11.24</b>		± 0.44			
HT	HCT	%	<b>16.0</b>	<b>15.9</b>	<b>15.9</b>		± 2.0	<b>36.5</b>	<b>36.4</b>	<b>36.8</b>		± 2.5	<b>48.8</b>	<b>49.0</b>	<b>49.3</b>		± 3.0			
		l/l	<b>0.160</b>	<b>0.159</b>	<b>0.159</b>		± 0.020	<b>0.365</b>	<b>0.364</b>	<b>0.368</b>		± 0.025	<b>0.488</b>	<b>0.490</b>	<b>0.493</b>		± 0.030			
VGM	MCV	µm <sup>3</sup> ; fl	<b>71</b>	<b>71</b>	<b>71.5</b>		± 4	<b>81</b>	<b>82</b>	<b>81.0</b>		± 4	<b>89</b>	<b>89</b>	<b>87.5</b>		± 4			
TGMH	MCH	pg	<b>26.2</b>	<b>26.3</b>	<b>25.6</b>		± 2.4	<b>29.6</b>	<b>30.0</b>	<b>29.7</b>		± 2.4	<b>32.1</b>	<b>32.4</b>	<b>32.1</b>		± 2.8			
		fmol	<b>1.63</b>	<b>1.64</b>	<b>1.59</b>		± 0.15	<b>1.84</b>	<b>1.86</b>	<b>1.85</b>		± 0.15	<b>1.99</b>	<b>2.01</b>	<b>2.00</b>		± 0.18			
CCMH	MCHC	g/dl	<b>36.9</b>	<b>37.1</b>	<b>35.7</b>		± 3.0	<b>36.5</b>	<b>36.5</b>	<b>36.7</b>		± 3.0	<b>36.1</b>	<b>36.4</b>	<b>36.7</b>		± 3.0			
		g/l	<b>369</b>	<b>371</b>	<b>357</b>		± 30	<b>365</b>	<b>365</b>	<b>367</b>		± 30	<b>361</b>	<b>364</b>	<b>367</b>		± 30			
		mmol/l	<b>22.94</b>	<b>23.04</b>	<b>22.20</b>		± 1.86	<b>22.66</b>	<b>22.69</b>	<b>22.80</b>		± 1.86	<b>22.41</b>	<b>22.58</b>	<b>22.82</b>		± 1.86			
IDR	RDW	%	<b>15.0</b>	<b>14.5</b>	<b>16.5</b>		± 3.0	<b>13.5</b>	<b>13.0</b>	<b>13.5</b>		± 3.0	<b>13.0</b>	<b>12.5</b>	<b>12.5</b>		± 3.0			
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>59</b>	<b>62</b>	<b>57</b>		± 20	<b>250</b>	<b>249</b>	<b>256</b>		± 40	<b>489</b>	<b>501</b>	<b>507</b>		± 60			
VPM	MPV	µm <sup>3</sup> ; fl	<b>7.6</b>	<b>8.3</b>	<b>8.9</b>		± 2.0	<b>7.7</b>	<b>7.7</b>	<b>8.3</b>		± 2.0	<b>8.1</b>	<b>8.2</b>	<b>8.7</b>		± 2.0			
LYMPHO		#	<b>1.4</b>	<b>1.4</b>	N/A		± 0.2	<b>2.9</b>	<b>2.8</b>	N/A		± 0.5	<b>3.5</b>	<b>3.3</b>	N/A		± 1.1			
		%	<b>66.5</b>	<b>65.5</b>	N/A		± 8.0	<b>37.0</b>	<b>36.5</b>	N/A		± 6.0	<b>17.0</b>	<b>16.0</b>	N/A		± 5.0			
MONO		#	<b>0.2</b>	<b>0.2</b>	N/A		± 0.2	<b>0.5</b>	<b>0.5</b>	N/A		± 0.4	<b>2.0</b>	<b>2.1</b>	N/A		± 1.0			
		%	<b>7.5</b>	<b>8.5</b>	N/A		± 7.0	<b>6.0</b>	<b>6.0</b>	N/A		± 5.0	<b>9.5</b>	<b>10.1</b>	N/A		± 5.0			
GRANULO		#	<b>0.5</b>	<b>0.5</b>	N/A		± 0.2	<b>4.5</b>	<b>4.4</b>	N/A		± 0.5	<b>15.2</b>	<b>15.1</b>	N/A		± 1.5			
		%	<b>26.0</b>	<b>26.0</b>	N/A		± 7.0	<b>57.0</b>	<b>57.5</b>	N/A		± 6.0	<b>73.5</b>	<b>73.9</b>	N/A		± 7.0			
CRP		mg/dl	<b>0.46</b>	<b>0.44</b>	<b>0.49</b>		± 0.40	<b>2.43</b>	<b>2.39</b>	<b>2.72</b>		± 0.80	<b>4.42</b>	<b>4.33</b>	<b>4.41</b>		± 1.50			
		mg/l	<b>4.6</b>	<b>4.4</b>	<b>4.9</b>		± 4.0	<b>24.3</b>	<b>23.9</b>	<b>27.2</b>		± 8.0	<b>44.2</b>	<b>43.3</b>	<b>44.1</b>		± 15.0			

Ref: TEMP-1070 Rev.42 BACK/VERSO 1300092170