

ABX Difftrol



LOT PX 436
Rev 1

CONTROL

(Exp.) **2022-09-05**
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Lysebio														
		CONTROL		L	TOLERANCES TOLERANCE	CONTROL		N	TOLERANCES TOLERANCE	CONTROL		H	TOLERANCES TOLERANCE			
		YUMIZEN		H1500		H2500	YUMIZEN			H1500	H2500	YUMIZEN		H1500	H2500	
		H1500	H2500				H1500	H2500				H1500				H2500
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.40	2.40		± 0.40	7.40	7.40		± 1.00	17.80	17.80		± 2.20			
GR RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.44	2.44		± 0.16	4.76	4.76		± 0.20	5.24	5.24		± 0.25			
HB HGB	g/dl	6.3	6.3		± 0.4	13.7	13.7		± 0.5	16.3	16.3		± 0.6			
	g/l	63	63		± 4	137	137		± 5	163	163		± 6			
	mmol/l	3.91	3.91		± 0.25	8.51	8.51		± 0.31	10.12	10.12		± 0.37			
HT HCT	%	19.2	19.2		± 1.5	41.1	41.1		± 2.0	48.8	48.8		± 2.5			
	l/l	0.192	0.192		± 0.015	0.411	0.411		± 0.020	0.488	0.488		± 0.025			
VGM MCV	µm ³ ; fl	78.6	78.6		± 5.0	86.6	86.6		± 5.0	93.5	93.5		± 5.0			
TGMH MCH	pg	25.8	25.8		± 2.0	28.8	28.8		± 2.0	31.1	31.1		± 2.5			
	fmol	1.60	1.60		± 0.12	1.79	1.79		± 0.12	1.93	1.93		± 0.16			
CCMH MCHC	g/dl	32.8	32.8		± 3.0	33.2	33.2		± 3.0	33.3	33.3		± 3.0			
	g/l	328	328		± 30	332	332		± 30	333	333		± 30			
	mmol/l	20.40	20.40		± 1.86	20.64	20.64		± 1.86	20.66	20.66		± 1.86			
IDR-CV RDW-CV	%	14.5	14.5		± 4.0	12.5	12.5		± 4.0	11.5	11.5		± 4.0			
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	71	71		± 20	257	257		± 30	512	512		± 50			
VPM MPV	µm ³ ; fl	9.7	9.7		± 2.0	12.7	12.7		± 2.0	11.6	11.6		± 2.0			
NEUT	#	1.21	1.21		± 0.35	3.94	3.94		± 0.90	12.87	12.87		± 1.90			
	%	50.6	50.6		± 10.0	53.3	53.3		± 10.0	72.3	72.3		± 10.0			
LYMPHO	#	0.68	0.68		± 0.33	2.39	2.39		± 0.70	2.47	2.47		± 1.50			
	%	28.3	28.3		± 12.0	32.3	32.3		± 8.0	13.9	13.9		± 8.0			
MONO	#	0.23	0.23		± 0.23	0.44	0.44		± 0.40	0.73	0.73		± 0.66			
	%	9.6	9.6		± 9.6	6.0	6.0		± 5.4	4.1	4.1		± 3.7			
EOS	#	0.18	0.18		± 0.18	0.33	0.33		± 0.33	0.94	0.94		± 0.94			
	%	7.4	7.4		± 7.4	4.4	4.4		± 4.4	5.3	5.3		± 5.3			
BASO	#	0.10	0.10		± 0.10	0.30	0.30		± 0.20	0.78	0.78		± 0.39			
	%	4.1	4.1		± 2.5	4.0	4.0		± 3.0	4.4	4.4		± 2.5			
IDR-SD RDW-SD	fl	43.0	43.0		± 4.0	41.0	41.0		± 4.0	41.5	41.5		± 4.0			
ERB NRBC	#	0.42	0.42		± 0.13	1.48	1.48		± 0.30	1.12	1.12		± 0.25			
	%	17.7	17.7		± 4.0	20.0	20.0		± 5.0	6.3	6.3		± 1.5			
PLTS OPT	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	70		± 30	N/A	220		± 40	N/A	451		± 50			

1300105209 BACK / VERSO Ref: TEMP-0821 Rev.46