


LOT PX 444
Rev 1

CONTROL

 (Exp.) 2024-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.45	2.45	± 0.40		7.30	7.30		± 1.00	17.80		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.39	± 0.16	4.64	4.64	± 0.20	5.16	5.16	± 0.25		
HB HGB	g/dl	6.3	6.3	± 0.4	13.6	13.6	± 0.5	16.2	16.2	± 0.6		
	g/l	63	63	± 4	136	136	± 5	162	162	± 6		
HT HCT	mmol/l	3.91	3.91	± 0.25	8.45	8.45	± 0.31	10.06	10.06	± 0.37		
	%	18.8	18.8	± 1.5	39.7	39.7	± 2.0	48.0	48.0	± 2.5		
VGM MCV	l/l	0.188	0.188	± 0.015	0.397	0.397	± 0.020	0.480	0.480	± 0.025		
	µm ³ ; fl	78.5	78.5	± 5.0	85.5	85.5	± 5.0	93.0	93.0	± 5.0		
TGMH MCH	pg	26.4	26.4	± 2.0	29.3	29.3	± 2.0	31.4	31.4	± 2.5		
	fmol	1.64	1.64	± 0.12	1.82	1.82	± 0.12	1.95	1.95	± 0.16		
CCMH MCHC	g/dl	33.6	33.6	± 3.0	34.3	34.3	± 3.0	33.8	33.8	± 3.0		
	g/l	336	336	± 30	343	343	± 30	338	338	± 30		
IDR-SD RDW-SD	mmol/l	20.85	20.85	± 1.86	21.29	21.29	± 1.86	20.96	20.96	± 1.86		
IDR-CV RDW-CV	fl	43.5	43.5	± 8.0	40.5	40.5	± 8.0	40.5	40.5	± 8.0		
	%	14.0	14.0	± 4.0	12.0	12.0	± 4.0	11.5	11.5	± 4.0		
PLA PLT	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72	± 20	235	235	± 30	492	492	± 50		
PLT-Ox	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	75	± 30	N/A	235	± 40	N/A	470	± 50		
VMP MPV	µm ³ ; fl	8.6	8.6	± 2.0	9.2	9.2	± 2.0	10.0	10.0	± 2.0		
NEU	10 ⁹ /mm ³ ; 10 ⁹ /l	1.28	1.28	± 0.35	4.03	4.03	± 0.90	12.87	12.87	± 1.90		
	%	52.3	52.3	± 10.0	55.2	55.2	± 10.0	72.3	72.3	± 10.0		
LYM	10 ⁹ /mm ³ ; 10 ⁹ /l	0.72	0.72	± 0.33	2.25	2.25	± 0.70	2.60	2.60	± 1.50		
	%	29.2	29.2	± 12.0	30.8	30.8	± 8.0	14.6	14.6	± 8.0		
MON	10 ⁹ /mm ³ ; 10 ⁹ /l	0.17	0.17	± 0.17	0.43	0.43	± 0.39	0.66	0.66	± 0.59		
	%	6.8	6.8	± 6.80	5.9	5.9	± 5.3	3.7	3.7	± 3.3		
EOS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.18	0.18	± 0.18	0.28	0.28	± 0.28	0.87	0.87	± 0.87		
	%	7.3	7.3	± 7.30	3.8	3.8	± 3.8	4.9	4.9	± 4.9		
BAS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.11	0.11	± 0.11	0.31	0.31	± 0.21	0.80	0.80	± 0.40		
	%	4.4	4.4	± 2.5	4.3	4.3	± 3.0	4.5	4.5	± 2.5		
ERB NRBC	10 ⁹ /mm ³ ; 10 ⁹ /l	0.42	0.42	± 0.30	1.32	1.32	± 0.45	1.16	1.16	± 0.35		
	%	17.3	17.3	± 8.0	18.1	18.1	± 8.0	6.5	6.5	± 2.0		

BACK / VERSO
Ref: TEMP-0821 Rev.49