

LOT PX 443
Rev 2

CONTROL

 (Exp.) 2023-11-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		H1500 V1.0 to V1.4	H2500 V1.0 to V1.4		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.45	2.45	± 0.40		7.20	7.20		± 1.00	17.90		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.36	2.36	± 0.16	4.69	4.69	± 0.20	5.31	5.31	± 0.25		
HB HGB	g/dl	5.9	5.9	± 0.4	13.7	13.7	± 0.5	16.8	16.8	± 0.6		
	g/l	59	59	± 4	137	137	± 5	168	168	± 6		
	mmol/l	3.66	3.66	± 0.25	8.51	8.51	± 0.31	10.43	10.43	± 0.37		
HT HCT	%	17.7	17.7	± 1.5	40.3	40.3	± 2.0	50.2	50.2	± 2.5		
	l/l	0.177	0.177	± 0.015	0.403	0.403	± 0.020	0.502	0.502	± 0.025		
VGM MCV	µm ³ ; fl	75.0	75.0	± 5.0	86.0	86.0	± 5.0	94.5	94.5	± 5.0		
TGMH MCH	pg	25.0	25.0	± 2.0	29.2	29.2	± 2.0	31.6	31.6	± 2.5		
	fmol	1.55	1.55	± 0.12	1.81	1.81	± 0.12	1.96	1.96	± 0.16		
CCMH MCHC	g/dl	33.3	33.3	± 3.0	34.0	34.0	± 3.0	33.5	33.5	± 3.0		
	g/l	333	333	± 30	340	340	± 30	335	335	± 30		
	mmol/l	20.70	20.70	± 1.86	21.09	21.09	± 1.86	20.79	20.79	± 1.86		
IDR-SD RDW-SD	fl	43.5	43.5	± 8.0	41.5	41.5	± 8.0	42.0	42.0	± 8.0		
IDR-CV RDW-CV	%	15.0	15.0	± 4.0	12.0	12.0	± 4.0	11.5	11.5	± 4.0		
PLA PLT	10 ⁹ /mm ³ ; 10 ⁹ /l	67	67	± 20	254	254	± 30	497	497	± 50		
PLT-Ox	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	69	± 30	N/A	245	± 40	N/A	477	± 50		
VMP MPV	µm ³ ; fl	9.5	9.5	± 2.0	9.8	9.8	± 2.0	11.0	11.0	± 2.0		
NEU	10 ⁹ /mm ³ ; 10 ⁹ /l	1.24	1.24	± 0.35	3.87	3.87	± 0.90	12.87	12.87	± 1.90		
	%	50.8	50.8	± 10.0	53.8	53.8	± 10.0	71.9	71.9	± 10.0		
LYM	10 ⁹ /mm ³ ; 10 ⁹ /l	0.75	0.75	± 0.33	2.23	2.23	± 0.70	2.51	2.51	± 1.50		
	%	30.5	30.5	± 12.0	31.0	31.0	± 8.0	14.0	14.0	± 8.0		
MON	10 ⁹ /mm ³ ; 10 ⁹ /l	0.21	0.21	± 0.21	0.50	0.50	± 0.45	0.93	0.93	± 0.84		
	%	8.5	8.5	± 8.50	7.0	7.0	± 6.3	5.2	5.2	± 4.7		
EOS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.15	0.15	± 0.15	0.29	0.29	± 0.29	0.79	0.79	± 0.79		
	%	6.2	6.2	± 6.20	4.0	4.0	± 4.0	4.4	4.4	± 4.4		
BAS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.10	0.10	± 0.10	0.30	0.30	± 0.20	0.81	0.81	± 0.41		
	%	4.0	4.0	± 2.5	4.2	4.2	± 3.0	4.5	4.5	± 2.5		
ERB NRBC	10 ⁹ /mm ³ ; 10 ⁹ /l	0.46	0.46	± 0.30	1.30	1.30	± 0.45	1.11	1.11	± 0.35		
	%	18.6	18.6	± 8.0	18.0	18.0	± 8.0	6.2	6.2	± 2.0		

BACK / VERSO
Ref: TEMP-0821 Rev.49