


LOT PX 445
Rev 1

CONTROL

 (Exp.) 2024-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	Lysebio									
			CONTROL		L	CONTROL		N	CONTROL		H	TOLERANCES TOLERANCE
			YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	YUMIZEN		TOLERANCES TOLERANCE	
			H1500 HELO 1.x	H2500 HELO 1.x		H1500 HELO 1.x	H2500 HELO 1.x		H1500 HELO 1.x	H2500 HELO 1.x		
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.45	2.45	± 0.40	7.40	7.40	± 1.00	17.90	17.90	± 2.20		
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.42	2.42	± 0.16	4.60	4.60	± 0.20	5.05	5.05	± 0.25		
HB HGB	g/dl	6.2	6.2	± 0.4	13.6	13.6	± 0.5	16.0	16.0	± 0.6		
	g/l	62	62	± 4	136	136	± 5	160	160	± 6		
	mmol/l	3.85	3.85	± 0.25	8.45	8.45	± 0.31	9.94	9.94	± 0.37		
HT HCT	%	18.6	18.6	± 1.5	40.0	40.0	± 2.0	47.5	47.5	± 2.5		
	l/l	0.186	0.186	± 0.015	0.400	0.400	± 0.020	0.475	0.475	± 0.025		
VGM MCV	µm ³ ; fl	77.0	77.0	± 5.0	87.0	87.0	± 5.0	94.0	94.0	± 5.0		
TGMH MCH	pg	25.6	25.6	± 2.0	29.6	29.6	± 2.0	31.7	31.7	± 2.5		
	fmol	1.59	1.59	± 0.12	1.84	1.84	± 0.12	1.97	1.97	± 0.16		
CCMH MCHC	g/dl	33.3	33.3	± 3.0	34.0	34.0	± 3.0	33.7	33.7	± 3.0		
	g/l	333	333	± 30	340	340	± 30	337	337	± 30		
	mmol/l	20.66	20.66	± 1.86	21.10	21.10	± 1.86	20.93	20.93	± 1.86		
IDR-SD RDW-SD	fl	43.5	43.5	± 8.0	41.0	41.0	± 8.0	40.5	40.5	± 8.0		
IDR-CV RDW-CV	%	15.0	15.0	± 4.0	13.0	13.0	± 4.0	11.0	11.0	± 4.0		
PLA PLT	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72	± 20	240	240	± 30	483	483	± 50		
PLT-Ox	10 ⁹ /mm ³ ; 10 ⁹ /l	N/A	75	± 30	N/A	235	± 40	N/A	462	± 50		
VMP MPV	µm ³ ; fl	9.4	9.4	± 2.0	9.6	9.6	± 2.0	10.2	10.2	± 2.0		
NEU	10 ⁹ /mm ³ ; 10 ⁹ /l	1.24	1.24	± 0.35	3.97	3.97	± 0.90	12.76	12.76	± 1.90		
	%	50.6	50.6	± 10.0	53.7	53.7	± 10.0	71.3	71.3	± 10.0		
LYM	10 ⁹ /mm ³ ; 10 ⁹ /l	0.75	0.75	± 0.33	2.44	2.44	± 0.70	2.76	2.76	± 1.50		
	%	30.8	30.8	± 12.0	33.0	33.0	± 8.0	15.4	15.4	± 8.0		
MON	10 ⁹ /mm ³ ; 10 ⁹ /l	0.20	0.20	± 0.20	0.43	0.43	± 0.39	0.77	0.77	± 0.69		
	%	8.3	8.3	± 8.30	5.8	5.8	± 5.2	4.3	4.3	± 3.9		
EOS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.15	0.15	± 0.15	0.26	0.26	± 0.26	0.81	0.81	± 0.81		
	%	6.1	6.1	± 6.10	3.5	3.5	± 3.5	4.5	4.5	± 4.5		
BAS	10 ⁹ /mm ³ ; 10 ⁹ /l	0.10	0.10	± 0.10	0.30	0.30	± 0.20	0.81	0.81	± 0.41		
	%	4.2	4.2	± 2.5	4.0	4.0	± 3.0	4.5	4.5	± 2.5		
ERB NRBC	10 ⁹ /mm ³ ; 10 ⁹ /l	0.41	0.41	± 0.30	1.45	1.45	± 0.45	1.24	1.24	± 0.35		
	%	18.2	18.2	± 8.0	19.6	19.6	± 8.0	6.9	6.9	± 2.0		

BACK / VERSO
Ref: TEMP-0821 Rev.50