

ABX Minotrol Retic



LOT RX 445
Rev 1

CONTROL

(Exp.) 2024-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
		H2500 HELO 1.x				H2500 HELO 1.x				H2500 HELO 1.x			
Méthode Manuelle / Manual													
RET	%	1.4			± 0.7	5.4			± 2.2	10.4			± 3.2
RET	%	1.43			± 0.72	5.02			± 1.26	10.16			± 2.03
RET	10 ⁶ /mm ³	0.050			± 0.025	0.175			± 0.044	0.353			± 0.071
	10 ⁹ /l	50			± 25	175			± 44	353			± 71
MFI MRI *	%	24.7			± 8.0	30.5			± 10.0	31.9			± 12.0
	-	0.247			± 0.080	0.305			± 0.100	0.319			± 0.120
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	3.47			± 0.30	3.49			± 0.30	3.47			± 0.30
PIC RANGE	-	19			± 8	19			± 8	19			± 8

* For research use only in Canada and USA

Ref: TEMP-0382 Rev.45 FRONT / RECTO 1300116193