

# ABX Minotrol Retic



**LOT** RX 444

**CONTROL**

(Exp.) **2024-01-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		YUMIZEN				YUMIZEN				YUMIZEN			
		H2500 Since V1.7				H2500 Since V1.7				H2500 Since V1.7			
Méthode Manuelle / Manual													
RET	%	<b>1.6</b>			± 0.8	<b>4.9</b>			± 2.0	<b>10.5</b>			± 3.2
RET	%	<b>1.70</b>			± 0.85	<b>5.10</b>			± 1.28	<b>10.12</b>			± 2.02
RET	10 <sup>6</sup> /mm <sup>3</sup>	<b>0.058</b>			± 0.029	<b>0.175</b>			± 0.044	<b>0.344</b>			± 0.069
	10 <sup>9</sup> /l	<b>58</b>			± 29	<b>175</b>			± 44	<b>344</b>			± 69
CRC	%	<b>1.04</b>			± 0.52	<b>3.31</b>			± 0.99	<b>6.25</b>			± 1.56
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>3.40</b>			± 0.30	<b>3.43</b>			± 0.30	<b>3.40</b>			± 0.30
IRF	--	<b>0.40</b>			± 0.24	<b>0.58</b>			± 0.35	<b>0.62</b>			± 0.37
RHCC	pg	<b>22.2</b>			± 4.4	<b>16.9</b>			± 3.4	<b>12.0</b>			± 3.0

Ref: TEMP-0382 Rev.44 1300116193 BACK / VERSO