

ABX Pentra

CO2 Control

 **HORIBA ABX SAS**
 Parc Euromédecine - Rue du Caducée
 B.P. 7290
 34184 Montpellier Cedex 4
 FRANCE
 www.horiba.com

    **LOT** 1: 24830  **REF** : A11A01650

04710802H
 **SAP** 1300023125  2019/01/31

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidenstervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U		→ ±3 SD ←	1 SD
CO2 (UV enzymatic)	●	●	●	mmol/L	20	16 - 24	1.3

ABX Pentra

CO2 Control

 **HORIBA ABX SAS**
 Parc Euromédecine - Rue du Caducée
 B.P. 7290
 34184 Montpellier Cedex 4
 FRANCE
 www.horiba.com

    **LOT** 1: 24830  **REF** : A11A01650

04710802H
 **SAP** 1300023125  2019/01/31

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidenstervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U		→ ±3 SD ←	1 SD
CO2 (UV enzymatic)	●	●	●	mmol/L	20	16 - 24	1.3