

Yumizen G PT Reco 10 for Yumizen G Line



Reference No.:		1300036376	
Lot No.:		980501	
Expiry date:		2020-05-31	
Coagulometers		ISI	MASTER CURVE
Optical	Yumizen G Line		%
		100	11,1
		50	16,1
		33	21,2
		25	26,0
	1,09	MNPT	11,1
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>For Yumizen G200, Yumizen G400 and Yumizen G400 DDi, the Min. Time must be set to 8 seconds in the PT setup. Please refer to the Yumizen G PT Reco 10 reagent notice.</p>			