



HORIBA ABX SAS
Parc Euromédecine - Rue du Caducée
B.P. 7290
34184 Montpellier Cedex 4
FRANCE
www.horiba.com



LOT 1: 25862



REF : A11A01650



2019/09/30

SAP 1300057823

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklarierter Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidensintervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U	$\oplus$	Confidence range	1 SD
CO2 (UV enzymatic)	•	•	•	mmol/L	20	16 - 24	1.3



HORIBA ABX SAS
Parc Euromédecine - Rue du Caducée
B.P. 7290
34184 Montpellier Cedex 4
FRANCE
www.horiba.com



LOT 1: 25862



REF : A11A01650



2019/09/30

SAP 1300057823

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklarierter Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tilldelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidensintervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U	$\oplus$	Confidence range	1 SD
CO2 (UV enzymatic)	•	•	•	mmol/L	20	16 - 24	1.3