



# Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 8 0 7 0 6 4 6 4 1 2 0 7 2 0 D



D 5 1 0 7 1 4 4 0 8 0 1 4 8 0 7 5 D



D 6 1 1 3 1 0 3 1 3 9 0 8 0 1 3 4 D



D 7 0 7 7 1 1 6 0 8 3 1 1 2 1 0 1 D



D 8 1 6 9 0 7 0 1 0 5 3 2 7 4 4 3 D



D 9 3 0 4 4 1 1 1 9 0 2 8 0 2 1 2 D



D 0 3 1 7 3 2 3 D

Reference No.:		1300036412	
Lot No:		980706_REV2	
Expiry date:		2020-07-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,7-14,4
	Prothrombin %	%	80-148
	Prothrombin INR	INR	0,75-1,13
PT Liq	Prothrombin Time	sec	10,3-13,9
	Prothrombin %	%	80-134
	Prothrombin INR	INR	0,77-1,16
PT R	Prothrombin Time	sec	8,3-11,2
	Prothrombin %	%	101-169
	Prothrombin INR	INR	0,70-1,05
APTT	Partial Thromboplastin Time	sec	32,7-44,3
APTT Liq	Partial Thromboplastin Time	sec	30,4-41,1
FIB Cl	Fibrinogen	g/L	1,9-2,8
TT	Thrombin Time	sec	21,2-31,7
AT	Antithrombin III	%	84-113



# Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 8 0 7 0 6 5 6 4 1 2 0 7 2 0 D



D 5 1 5 7 2 3 6 0 3 3 0 6 2 1 2 7 D



D 6 1 9 0 1 4 8 2 2 2 0 3 4 0 5 7 D



D 7 1 3 1 1 9 6 1 1 3 1 7 0 0 4 7 D



D 8 0 7 8 1 0 9 1 6 4 5 4 6 7 3 9 D



D 9 4 9 7 6 7 2 0 7 0 1 7 0 0 0 0 D



D 0 9 9 9 3 8 4 D

Reference No.:		1300036412	
Lot No:		980706_REV2	
Expiry date:		2020-07-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	15,7-23,6
	Prothrombin %	%	33-62
	Prothrombin INR	INR	1,27-1,90
PT Liq	Prothrombin Time	sec	14,8-22,2
	Prothrombin %	%	34-57
	Prothrombin INR	INR	1,31-1,96
PT R	Prothrombin Time	sec	11,3-17,0
	Prothrombin %	%	47-78
	Prothrombin INR	INR	1,09-1,64
APTT	Partial Thromboplastin Time	sec	54,6-73,9
APTT Liq	Partial Thromboplastin Time	sec	49,7-67,2
FIB Cl	Fibrinogen	g/L	0,7-1,7
TT	Thrombin Time	sec	NA
AT	Antithrombin III	%	41-68