



# Yumizen G PT 5 for Yumizen G Line



A 4 9 9 0 2 2 1 3 6 3 3 8 0 2 2 1 A



A 5 1 0 0 1 3 0 5 0 0 1 8 9 3 3 0 A



A 6 2 5 0 2 5 0 3 0 7 1 3 0 1 2 2 A



A 7 1 4 0 A

|  |                       |                   |                     |             |
|--|-----------------------|-------------------|---------------------|-------------|
| <b>Reference No.:</b>  |                       | <b>1300036338</b> |                     |             |
| <b>Lot No.:</b>  |                       | <b>990221</b>     |                     |             |
| <b>Expiry date:</b>  |                       | <b>2021-02-28</b> |                     |             |
| <b>Coagulometers</b>   |                       | <b>ISI</b>        | <b>MASTER CURVE</b> |             |
| <b>Optical</b>   | <b>Yumizen G Line</b> |                   | <b>1,22</b>         | <b>%</b>    |
|  |                       | <b>100</b>        |                     | <b>13,0</b> |
|  |                       | <b>50</b>         |                     | <b>18,9</b> |
|  |                       | <b>33</b>         |                     | <b>25,0</b> |
|  |                       | <b>25</b>         |                     | <b>30,7</b> |
|  | <b>MNPT</b>           | <b>13,0</b>       |                     |             |
| <b>Important notice</b>  |                       |                   |                     |             |
| <ul style="list-style-type: none"> <li>- The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</li> <li>- The efficiency of the reconstitution of the reagent may be approved by performing it at 37°C.</li> </ul> |                       |                   |                     |             |