



HORIBA ABX SAS
 Parc Euromédecine - Rue du Caducée
 B.P. 7290
 34184 Montpellier Cedex 4
 FRANCE
 www.horiba.com



LOT 1: 1906501



1 9 0 6 5 0 1

REF : 1300026187



2022/06/30

SAP 1300056305



1300056305

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400	Pentra C400	Pentra C200	Unité	Valeur titrée	Intervalle de confiance	
				Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tildelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidensintervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U	\oplus	Confidence range*	1 SD
CK (IFCC/DGKC) (37°C)	●			U/L μkat/L	149 2.49	134 - 164 2,25 - 2,73	5 0.08
		●		U/L μkat/L	149 2.49	134 - 164 2,25 - 2,73	5 0.08
			●	U/L μkat/L	152 2.54	137 - 167 2,30 - 2,78	5 0.08
CK-MB (Immunology method) (37°C)	●			U/L μkat/L	32.0 0.534	28,7 - 35,3 0,480 - 0,588	1.1 0.018
		●		U/L μkat/L	32.0 0.534	28,7 - 35,3 0,480 - 0,588	1.1 0.018
			●	U/L μkat/L	40.9 0.683	36,7 - 45,1 0,614 - 0,752	1.4 0.023

* Confidence range: +/- 2 SD