



Yumizen G PT Reco 10 for Yumizen G Line



Reference No.:		1300036376		
Lot No.:		980715		
Expiry date:		2020-07-31		
Coagulometers		ISI	ISI	
Optical	Yumizen G Line	1,03	100	10,8
			50	15,6
			33	20,6
			25	25,2
			MNPT	10,8
Important notice				
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>For Yumizen G200, Yumizen G400 and Yumizen G400 DDi, the Min. Time must be set to 8 seconds in the PT setup. Please refer to the Yumizen G PT Reco 10 reagent notice.</p>				