



Yumizen G TT for Yumizen G Line



D 4 9 9 0 2 1 2 3 6 3 8 2 0 2 2 1 D



D 5 1 8 9 0 7 7 D

Reference No.:		1300036382
Lot No.:		990212
Expiry date:		2021-02-28
MNTT		
Coagulometers		sec
Optical	Yumizen G Line	18,9
Important notice		
<p>The Mean Normal Thrombin Time (MNTT) is for information only. The value depends on the population, race, gender, sampling tube etc. According to the CLSI every laboratory should determine its local MNTT.</p>		