



HORIBA ABX SAS
BP 7290
34184 Montpellier
cedex 4 - France
www.horiba-
abx.com



A03A00012A

Yumizen G APTT Liq 4 for YUMIZEN G Line



D 4 9 9 1 2 1 1 3 6 3 8 1 0 6 2 1 D



D 5 2 5 5 0 7 4 D

Reference No.:		1300036381
Lot No.:		991211
Expiry date:		2021-06-30
MNAPTT		
Coagulometers		Sec
Optical	Yumizen G Line	25,5
Important notice		
The Mean Normal Activated Partial Thromboplastin Time (MNAPTT) is for information only. The value depends on the population, race, gender, sampling tube, etc. According to the CLSI every laboratory should determine its local MNAPTT.		