

MINOCAL™ CALIBRATOR

Hematology Reference Calibrator

Instrument:

Lot No.: **CX446**

Performed by:

Exp. Date: **2020-06-05** (Y / M / D)

Date:

Table of Assigned Values	WBC K/μL	RBC M/μL	HGB g/dL	HCT %	PLT K/μL	MPV fL
ABX MICROS ESV 60	8.7 ± 0.2	4.53 ± 0.08	12.9 ± 0.2	37.4 ± 1.0	243 ± 12	9.3 ± 0.5
ABX MICROS 60, ADVIA® 60	9.0 ± 0.2	4.52 ± 0.08	13.1 ± 0.2	37.5 ± 1.0	246 ± 12	7.9 ± 0.5
ABX MICROS CRP 200	8.9 ± 0.2	4.50 ± 0.08	13.1 ± 0.2	37.1 ± 1.0	244 ± 12	8.6 ± 0.5
ABX PENTRA 60 OT, ABX PENTRA 60 C+	9.1 ± 0.2	4.55 ± 0.08	13.1 ± 0.2	36.2 ± 1.0	245 ± 12	9.7 ± 0.5
ABX PENTRA 80, ABX PENTRA XL80	9.0 ± 0.2	4.53 ± 0.08	13.1 ± 0.2	36.0 ± 1.0	245 ± 12	9.8 ± 0.5
ABX PENTRA XLR	9.0 ± 0.2	4.53 ± 0.08	13.0 ± 0.2	35.8 ± 1.0	241 ± 12	10.5 ± 0.5
ABX MICROS ES 60	9.2 ± 0.2	4.56 ± 0.08	13.2 ± 0.2	37.8 ± 1.0	253 ± 12	8.5 ± 0.5
ABX PENTRA DX 120, ABX PENTRA DF 120	9.3 ± 0.2	4.58 ± 0.06	13.1 ± 0.2	36.2 ± 1.0	250 ± 10	N/A
ABX PENTRA DX NEXUS, ABX PENTRA DF NEXUS	9.3 ± 0.2	4.58 ± 0.06	13.1 ± 0.2	36.4 ± 1.0	250 ± 10	N/A

Table 1. Determine Mean Value	WBC K/μL	RBC M/μL	HGB g/dL	HCT %	PLT K/μL
Sample 1					
2					
3					
4					
5					
6					
Total					
Mean Value (Total ÷ 6)					
SD					
%CV (SD ÷ Mean Value × 100)					

Table 2. Determine Calibration Factor	WBC K/μL	RBC M/μL	HGB g/dL	HCT %	PLT K/μL
<u>MINOCAL Assigned Value</u>	—	—	—	—	—
Mean Value	—	—	—	—	—
Calibration Factor (round off to 3 decimal places)					
MINOCAL Reading (Table 1, Sample 6)					
Corrected Calibrator Reading for Adjustment					

Table 3. Verify Calibration	WBC K/μL	RBC M/μL	HGB g/dL	HCT %	PLT K/μL
Sample 1					
2					
3					
Total					
Mean Value (Total ÷ 3)					

For technical assistance, call the Customer Support Center at 1-888-903-5001.