

# MINOCAL™ CALIBRATOR

## Hematology Reference Calibrator

Instrument:  
Performed by:  
Date:

Lot No.: **CX447**  
Exp. Date: **2020-07-05** (Y / M / D)

Table of Assigned Values	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT %	PLT K/ $\mu$ L	MPV fL
ABX MICROS ESV 60	9.0 ± 0.2	4.45 ± 0.08	13.1 ± 0.2	37.8 ± 1.0	246 ± 12	8.6 ± 0.5
ABX MICROS 60, ADVIA® 60	9.3 ± 0.2	4.45 ± 0.08	13.4 ± 0.2	38.0 ± 1.0	248 ± 12	7.2 ± 0.5
ABX MICROS CRP 200	9.2 ± 0.2	4.46 ± 0.08	13.4 ± 0.2	37.9 ± 1.0	247 ± 12	7.7 ± 0.5
ABX PENTRA 60 OT, ABX PENTRA 60 C+	9.4 ± 0.2	4.48 ± 0.08	13.3 ± 0.2	36.5 ± 1.0	250 ± 12	9.1 ± 0.5
ABX PENTRA 80, ABX PENTRA XL80	9.3 ± 0.2	4.46 ± 0.08	13.4 ± 0.2	36.7 ± 1.0	247 ± 12	9.1 ± 0.5
ABX PENTRA XLR	9.3 ± 0.2	4.46 ± 0.08	13.3 ± 0.2	36.5 ± 1.0	245 ± 12	9.6 ± 0.5
ABX MICROS ES 60	9.4 ± 0.2	4.48 ± 0.08	13.4 ± 0.2	37.9 ± 1.0	247 ± 12	8.0 ± 0.5
ABX PENTRA DX 120, ABX PENTRA DF 120	9.6 ± 0.2	4.49 ± 0.06	13.3 ± 0.2	36.5 ± 1.0	248 ± 10	N/A
ABX PENTRA DX NEXUS, ABX PENTRA DF NEXUS	9.6 ± 0.2	4.49 ± 0.06	13.3 ± 0.2	36.6 ± 1.0	248 ± 10	N/A

Table 1. Determine Mean Value	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT %	PLT K/ $\mu$ L
Sample 1					
2					
3					
4					
5					
6					
Total					
Mean Value (Total ÷ 6)					
SD					
%CV (SD ÷ Mean Value × 100)					

Table 2. Determine Calibration Factor	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT %	PLT K/ $\mu$ L
MINOCAL Assigned Value Mean Value	_____	_____	_____	_____	_____
Calibration Factor (round off to 3 decimal places)					
MINOCAL Reading (Table 1, Sample 6)					
Corrected Calibrator Reading for Adjustment					

Table 3. Verify Calibration	WBC K/ $\mu$ L	RBC M/ $\mu$ L	HGB g/dL	HCT %	PLT K/ $\mu$ L
Sample 1					
2					
3					
Total					
Mean Value (Total ÷ 3)					

For technical assistance, call the Customer Support Center at 1-888-903-5001.