



Yumizen G PT Liq 4 for Yumizen G Line



A 4 9 0 0 2 1 2 3 6 3 7 3 0 8 2 1 A



A 5 1 0 0 1 3 4 5 0 0 1 9 4 3 3 0 A



A 6 2 5 7 2 5 0 3 1 5 1 3 4 1 2 8 A



A 7 1 5 2 A

Reference No.:		1300036373		
Lot No.:		900212		
Expiry date:		2021-08-31		
Coagulometers		ISI	MASTER CURVE	
Optical	Yumizen G Line		1,28	%
		100		13,4
		50		19,4
		33		25,7
		25		31,5
	MNPT	13,4		
Important notice				
The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.				