

MINOCAL™ CALIBRATOR

Hematology Reference Calibrator

Instrument:

Lot No.: **CX448**

Performed by:

Exp. Date: **2020-08-05** (Y / M / D)

Date:

| Table of Assigned Values | WBC K/μL | RBC M/μL | HGB g/dL | HCT % | PLT K/μL | MPV fL |
|--|--------------------------------|--------------------------------|-----------------|--------------|--------------------------------|---------------|
| ABX MICROS ESV 60 | 8.9 ± 0.2 | 4.35 ± 0.08 | 13.4 ± 0.2 | 38.5 ± 1.0 | 250 ± 12 | 8.9 ± 0.5 |
| ABX MICROS 60, ADVIA® 60 | 9.1 ± 0.2 | 4.35 ± 0.08 | 13.6 ± 0.2 | 38.5 ± 1.0 | 245 ± 12 | 7.3 ± 0.5 |
| ABX MICROS CRP 200 | 9.1 ± 0.2 | 4.38 ± 0.08 | 13.8 ± 0.2 | 38.8 ± 1.0 | 244 ± 12 | 8.0 ± 0.5 |
| ABX PENTRA 60 OT, ABX PENTRA 60 C+ | 9.2 ± 0.2 | 4.37 ± 0.08 | 13.6 ± 0.2 | 37.2 ± 1.0 | 246 ± 12 | 9.1 ± 0.5 |
| ABX PENTRA 80, ABX PENTRA XL80 | 9.1 ± 0.2 | 4.35 ± 0.08 | 13.6 ± 0.2 | 37.2 ± 1.0 | 247 ± 12 | 9.2 ± 0.5 |
| ABX PENTRA XLR | 9.1 ± 0.2 | 4.35 ± 0.08 | 13.5 ± 0.2 | 37.0 ± 1.0 | 237 ± 12 | 9.6 ± 0.5 |
| ABX MICROS ES 60 | 9.1 ± 0.2 | 4.34 ± 0.08 | 13.5 ± 0.2 | 38.2 ± 1.0 | 250 ± 12 | 7.9 ± 0.5 |
| ABX PENTRA DX 120, ABX PENTRA DF 120 | 9.5 ± 0.2 | 4.39 ± 0.06 | 13.6 ± 0.2 | 37.1 ± 1.0 | 249 ± 10 | N/A |
| ABX PENTRA DX NEXUS, ABX PENTRA DF NEXUS | 9.5 ± 0.2 | 4.39 ± 0.06 | 13.6 ± 0.2 | 37.3 ± 1.0 | 249 ± 10 | N/A |

| Table 1. Determine Mean Value | WBC K/μL | RBC M/μL | HGB g/dL | HCT % | PLT K/μL |
|--------------------------------------|--------------------------------|--------------------------------|-----------------|--------------|--------------------------------|
| Sample 1 | | | | | |
| 2 | | | | | |
| 3 | | | | | |
| 4 | | | | | |
| 5 | | | | | |
| 6 | | | | | |
| Total | | | | | |
| Mean Value (Total ÷ 6) | | | | | |
| SD | | | | | |
| %CV (SD ÷ Mean Value × 100) | | | | | |

| Table 2. Determine Calibration Factor | WBC K/μL | RBC M/μL | HGB g/dL | HCT % | PLT K/μL |
|--|--------------------------------|--------------------------------|-----------------|--------------|--------------------------------|
| <u>MINOCAL Assigned Value</u> | — | — | — | — | — |
| Mean Value | | | | | |
| Calibration Factor (round off to 3 decimal places) | | | | | |
| MINOCAL Reading (Table 1, Sample 6) | | | | | |
| Corrected Calibrator Reading for Adjustment | | | | | |

| Table 3. Verify Calibration | WBC K/μL | RBC M/μL | HGB g/dL | HCT % | PLT K/μL |
|------------------------------------|--------------------------------|--------------------------------|-----------------|--------------|--------------------------------|
| Sample 1 | | | | | |
| 2 | | | | | |
| 3 | | | | | |
| Total | | | | | |
| Mean Value (Total ÷ 3) | | | | | |

For technical assistance, call the Customer Support Center at 1-888-903-5001.