



# Yumizen G DDi 2 for YUMIZEN G1500/G1550



D 4991219363921220D



D 5400268030020602D



D 6001250100053505D



D 7002600250140161D

Reference No.:		1300036391	
Lot No.:		991219 (rev. 2)	
Expiry date:		2020-12-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G1500/G1550	$\mu\text{g(FEU)/mL}$	$\Delta\text{OD}$
		4,00	0,2680
		3,00	0,2060
		2,00	0,1250
		1,00	0,0535
		0,50	0,0260
		0,25	0,0140
		Slope value	0,06

# Yumizen G DDi 2 for YUMIZEN G800



D 4991219363921220D



D 5400243530018102D



D 6001080100047005D



D 7002300250125160D

Reference No.:		1300036391	
Lot No.:		991219 (rev. 2)	
Expiry date:		2020-12-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G800	$\mu\text{g(FEU)/mL}$	$\Delta\text{OD}$
		4,00	0,2435
		3,00	0,1810
		2,00	0,1080
		1,00	0,0470
		0,50	0,0230
		0,25	0,0125
		Slope value	0,06



# Yumizen G DDi 2 for YUMIZEN G400 DDi/G200

<b>Reference No.:</b>		<b>1300036391</b>	
<b>Lot No.:</b>		<b>991219 (rev. 2)</b>	
<b>Expiry date:</b>		<b>2020-12-31</b>	
<b>Coagulometers</b>		<b>MASTER CURVE</b>	
		<b>µg(FEU)/mL</b>	<b>ΔOD</b>
<b>Optical</b>	<b>Yumizen G400 DDi/G200</b>	<b>4,00</b>	<b>0,1015</b>
		<b>3,00</b>	<b>0,0743</b>
		<b>2,00</b>	<b>0,0480</b>
		<b>1,00</b>	<b>0,0231</b>
		<b>0,50</b>	<b>0,0115</b>
		<b>0,25</b>	<b>0,0057</b>
		<b>Slope value</b>	<b>0,04</b>
<b>Important notice</b>			
<p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type <b>Lin-Lin p-p</b> is selected in <b>Calibration &gt; D-Dimer &gt; ug FEU/ml</b>.          For more information, please contact your local HORIBA Medical representative.</p>			