



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 9 0 7 3 2 4 6 4 1 2 0 7 2 1 D



D 5 1 1 1 1 5 0 0 7 0 1 1 7 0 8 4 D



D 6 1 2 6 1 1 2 1 5 1 0 7 1 1 1 8 D



D 7 0 8 3 1 2 4 0 8 1 1 1 0 0 8 7 D



D 8 1 4 5 0 7 6 1 1 4 3 1 3 4 2 4 D



D 9 3 2 3 4 3 7 2 0 0 3 0 0 2 1 3 D



D 0 3 2 0 2 9 9 D

Reference No.:		1300036412	
Lot No:		990732 (rev. 3)	
Expiry date:		2021-07-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	11,1-15,0
	Prothrombin %	%	70-117
	Prothrombin INR	INR	0,84-1,26
PT Liq	Prothrombin Time	sec	11,2-15,1
	Prothrombin %	%	71-118
	Prothrombin INR	INR	0,83-1,24
PT R	Prothrombin Time	sec	8,1-11,0
	Prothrombin %	%	87-145
	Prothrombin INR	INR	0,76-1,14
APTT	Partial Thromboplastin Time	sec	31,3-42,4
APTT Liq (**)	Partial Thromboplastin Time	sec	32,3-43,7 (**)
FIB Cl	Fibrinogen	g/L	2,0-3,0
TT	Thrombin Time	sec	21,3-32,0
AT	Antithrombin	%	80-108
Important notice			
The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.			
The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.			

(**):Warning! This range has to be used only with APTTLiq reagents Lots n°: 900412 and higher (90xxxx where xxxx>0412).
For other lots, please refer to the previous revision of values sheet.



Yumizen G CTRL II (level 2) for Yumizen G Line



Reference No.:		1300036412	
Lot No:		990732 (rev. 3)	
Expiry date:		2021-07-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	16,1-24,1
	Prothrombin %	%	30-51
	Prothrombin INR	INR	1,42-2,12
PT Liq	Prothrombin Time	sec	15,4-23,1
	Prothrombin %	%	36-60
	Prothrombin INR	INR	1,33-2,00
PT R	Prothrombin Time	sec	11,3-16,9
	Prothrombin %	%	45-75
	Prothrombin INR	INR	1,07-1,61
APTT	Partial Thromboplastin Time	sec	47,8-64,6
APTT Liq (**)	Partial Thromboplastin Time	sec	45,8-62,0 (**)
FIB Cl	Fibrinogen	g/L	0,9-2,0
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	40-66

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(**): Warning! This range has to be used only with APTTLiq reagents Lots n°: 900412 and higher (90xxxx where xxxx>0412).

For other lots, please refer to the previous revision of values sheet.