



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 9 1 1 1 2 4 6 4 1 2 1 1 2 1 D



D 5 1 0 3 1 3 9 0 8 1 1 3 5 0 7 7 D



D 6 1 1 6 1 1 0 1 4 9 0 7 4 1 2 3 D



D 7 0 8 1 1 2 1 0 8 5 1 1 4 0 8 0 D



D 8 1 3 3 0 7 8 1 1 7 3 4 1 4 6 2 D



D 9 2 6 7 3 6 1 2 1 0 3 1 0 1 8 9 D



D 0 2 8 4 3 2 9 D

Reference No.:		1300036412	
Lot No:		991112	
Expiry date:		2021-11-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,3-13,9
	Prothrombin %	%	81-135
	Prothrombin INR	INR	0,77-1,16
PT Liq	Prothrombin Time	sec	11,0-14,9
	Prothrombin %	%	74-123
	Prothrombin INR	INR	0,81-1,21
PT R	Prothrombin Time	sec	8,5-11,4
	Prothrombin %	%	80-133
	Prothrombin INR	INR	0,78-1,17
APTT	Partial Thromboplastin Time	sec	34,1-46,2
APTT Liq	Partial Thromboplastin Time	sec	26,7-36,1
FIB Cl	Fibrinogen	g/L	2,1-3,1
TT	Thrombin Time	sec	18,9-28,4
AT	Antithrombin III	%	86-117



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 9 1 1 1 2 5 6 4 1 2 1 1 2 1 D



D 5 1 3 9 2 0 8 0 3 7 0 6 2 1 1 9 D



D 6 1 7 8 1 5 2 2 2 8 0 3 6 0 6 1 D



D 7 1 3 3 1 9 9 1 1 6 1 7 4 0 4 0 D



D 8 0 6 6 1 1 3 1 6 9 4 5 9 6 2 1 D



D 9 3 7 2 5 0 4 0 7 0 1 7 0 0 0 0 D



D 0 9 9 9 3 6 2 D

Reference No.:		1300036412	
Lot No:		991112	
Expiry date:		2021-11-30	
Optical Kit	Yumizen G Line Parameter	Unit	Range
PT	Prothrombin Time	sec	13,9-20,8
	Prothrombin %	%	37-62
	Prothrombin INR	INR	1,19-1,78
PT Liq	Prothrombin Time	sec	15,2-22,8
	Prothrombin %	%	36-61
	Prothrombin INR	INR	1,33-1,99
PT R	Prothrombin Time	sec	11,6-17,4
	Prothrombin %	%	40-66
	Prothrombin INR	INR	1,13-1,69
APTT	Partial Thromboplastin Time	sec	45,9-62,1
APTT Liq	Partial Thromboplastin Time	sec	37,2-50,4
FIB Cl	Fibrinogen	g/L	0,7-1,7
TT	Thrombin Time	sec	N/A
AT	Antithrombin III	%	40-66