



Yumizen G CTRL I (level 1) for Yumizen G Line



Reference No.:		1300036412	
Lot No:		991113	
Expiry date:		2021-11-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,3-13,9
	Prothrombin %	%	81-135
	Prothrombin INR	INR	0,77-1,16
PT Liq	Prothrombin Time	sec	11,0-14,9
	Prothrombin %	%	74-123
	Prothrombin INR	INR	0,81-1,21
PT R	Prothrombin Time	sec	8,5-11,4
	Prothrombin %	%	80-133
	Prothrombin INR	INR	0,78-1,17
APTT	Partial Thromboplastin Time	sec	34,1-46,2
APTT Liq	Partial Thromboplastin Time	sec	26,7-36,1
FIB Cl	Fibrinogen	g/L	2,1-3,1
TT	Thrombin Time	sec	18,9-28,4
AT	Antithrombin	%	86-117
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			



Yumizen G CTRL II (level 2) for Yumizen G Line



D4991113564121121D



D5150225034056130D



D6195155233035059D



D7136204118177036D



D8060119178511692D



D941055508019000D



D0999351D

Reference No.:		1300036412	
Lot No:		991113	
Expiry date:		2021-11-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	15,0-22,5
	Prothrombin %	%	34-56
	Prothrombin INR	INR	1,30-1,95
PT Liq	Prothrombin Time	sec	15,5-23,3
	Prothrombin %	%	35-59
	Prothrombin INR	INR	1,36-2,04
PT R	Prothrombin Time	sec	11,8-17,7
	Prothrombin %	%	36-60
	Prothrombin INR	INR	1,19-1,78
APTT	Partial Thromboplastin Time	sec	51,1-69,2
APTT Liq	Partial Thromboplastin Time	sec	41,0-55,5
FIB Cl	Fibrinogen	g/L	0,8-1,9
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	42-70
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			