



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 9 1 2 3 9 4 6 4 1 2 1 2 2 1 D



D 5 1 1 1 1 5 0 6 9 1 1 5 0 8 4 D



D 6 1 2 6 1 1 5 1 5 6 0 6 7 1 1 2 D



D 7 0 8 6 1 2 9 0 9 0 1 2 1 0 7 5 D



D 8 1 2 4 0 8 0 1 2 1 3 6 4 4 9 2 D



D 9 2 8 2 3 8 1 1 8 0 2 7 0 2 0 6 D



D 0 3 0 9 3 4 3 D

Reference No.:		1300036412	
Lot No:		991239 (rev 1.)	
Expiry date:		2021-12-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	11,1-15,0
	Prothrombin %	%	69-115
	Prothrombin INR	INR	0,84-1,26
PT Liq	Prothrombin Time	sec	11,5-15,6
	Prothrombin %	%	67-112
	Prothrombin INR	INR	0,86-1,29
PT R	Prothrombin Time	sec	9,0-12,1
	Prothrombin %	%	75-124
	Prothrombin INR	INR	0,80-1,21
APTT	Partial Thromboplastin Time	sec	36,4-49,2
APTT Liq	Partial Thromboplastin Time	sec	28,2-38,1
FIB Cl	Fibrinogen	g/L	1,8-2,7
TT	Thrombin Time	sec	20,6-30,9
AT	Antithrombin	%	88-119

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 9 1 2 3 9 5 6 4 1 2 1 2 2 1 D



D 5 1 5 0 2 2 5 0 3 4 0 5 7 1 2 6 D



D 6 1 8 8 1 6 1 2 4 1 0 3 3 0 5 6 D



D 7 1 4 2 2 1 2 1 1 7 1 7 6 0 3 9 D



D 8 0 6 5 1 1 4 1 7 1 5 2 1 7 0 5 D



D 9 4 2 2 5 7 1 0 7 0 1 7 0 0 0 D



D 0 9 9 9 3 4 4 D

Reference No.:		1300036412	
Lot No:		991239 (rev 1.)	
Expiry date:		2021-12-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	15,0-22,5
	Prothrombin %	%	34-57
	Prothrombin INR	INR	1,26-1,88
PT Liq	Prothrombin Time	sec	16,1-24,1
	Prothrombin %	%	33-56
	Prothrombin INR	INR	1,42-2,12
PT R	Prothrombin Time	sec	11,7-17,6
	Prothrombin %	%	39-65
	Prothrombin INR	INR	1,14-1,71
APTT	Partial Thromboplastin Time	sec	52,1-70,5
APTT Liq	Partial Thromboplastin Time	sec	42,2-57,1
FIB Cl	Fibrinogen	g/L	0,7-1,7
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	39-65
Important notice			
The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.			
The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.			