HORIBA ABX SAS
BP 7290
34184 Montpellier
cedex 4 – France
www.horiba-abx.com/documentation

IVD

A03A00001A

Yumizen G APTT 4
for Yumizen G Line

D490080636377 0822D

D5347084D

Reference No.:		1300036377
Lot No.:		900806 (rev. 1)
Expiry date:		2022-08-31
MNAPTT		
Coagulometers		Sec
Optical	Yumizen G Line	34,7

Important notice

The Mean Normal Activated Partial Thromboplastin Time (MNAPTT) is for information only. The value depends on the population, race, gender, sampling tube, etc. According to the CLSI every laboratory should determine its local MNAPTT.

Explore the future Automotive Test Systems I Process & Environmental I Medical I Semiconductor I Scientific HORIBA