

Yumizen G DDi 2 for YUMIZEN G1500/G1550



Reference No.:		1300036391	
Lot No.:		901209 (rev. 1)	
Expiry date:		2021-12-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G1500/G1550	µg(FEU)/mL	ΔOD
		4,00	0,2549
		3,00	0,1965
		2,00	0,1260
		1,00	0,0575
		0,50	0,0285
		0,25	0,0152
Slope value		0,07	

Yumizen G DDi 2 for YUMIZEN G800



Reference No.:		1300036391	
Lot No.:		901209 (rev. 1)	
Expiry date:		2021-12-31	
Coagulometers		MASTER CURVE	
Optical	Yumizen G800	µg(FEU)/mL	ΔOD
		4,00	0,2285
		3,00	0,1730
		2,00	0,1095
		1,00	0,0505
		0,50	0,0255
		0,25	0,0140
Slope value		0,07	



Yumizen G DDi 2 for YUMIZEN G400 DDi/G200

Reference No.:		1300036391	
Lot No.:		901209 (rev. 1)	
Expiry date:		2021-12-31	
Coagulometers		MASTER CURVE	
		µg(FEU)/mL	ΔOD
Optical	Yumizen G400 DDi/G200	4,00	0,1070
		3,00	0,0824
		2,00	0,0543
		1,00	0,0250
		0,50	0,0125
		0,25	0,0075
		Slope value	N/A
Important notice			
<p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type Lin-Lin p-p is selected in Calibration > D-Dimer > ug FEU/ml. For more information, please contact your local HORIBA Medical representative.</p>			