

**HORIBA ABX SAS**  
 Parc Euromédecine  
 Rue du Caducée  
 B.P. 7290  
 34184 Montpellier Cedex 4  
 FRANCE  
 www.horiba.com



**LOT** 21021 (rev. 01)



2 1 0 2 1

**SAP** 1300029568



1300029568

**REF** : 1300023947



2021/12/31

### For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400 (P400)	Pentra C400 (PC400)	Pentra C200 (PC200)	Unité Unit Einheit Unidad Unità Unidade Enhed Enhet Μονάδα Jednostka Enhet	Valeur titrée Assigned value Deklariertes Wert Valor asignado Valori assegnati Valor determinado Tildelt værdi Tildelat värde Αναγραφόμενη τιμή Wartość przypisana Tildelt verd	Intervalle de confiance Confidence range Vertrauensintervalle Intervalos de confianza Livelli di fiducia Intervalo de confiança Konfidensområde Konfidensintervall Εύρος εμπιστοσύνης Zakres ufności Konfidensspekter	
C				U	$\oplus$	Confidence range* 1 SD	
AlbT2-U ABX Pentra Micro ALBUMIN 2 CP (1300032563)	●			mg/L mg/dL	<b>65.4</b> <b>6.54</b>	52.4 – 78.4 5.24 – 7.84	6.5 0.65
		●		mg/L mg/dL	<b>65.4</b> <b>6.54</b>	52.4 – 78.4 5.24 – 7.84	6.5 0.65
			●	mg/L mg/dL	<b>62.7</b> <b>6.27</b>	50.1 – 75.3 5.01 – 7.53	6.3 0.63
AMY (EPS-G7) ABX Pentra Amylase CP (A11A01628)	●			U/L μkat/L	<b>243.0</b> <b>4.058</b>	194.4 - 291.6 3.246 - 4.870	24.3 0.406
		●		U/L μkat/L	<b>243.0</b> <b>4.058</b>	194.4 - 291.6 3.246 - 4.870	24.3 0.406
			●	U/L μkat/L	<b>234.7</b> <b>3.919</b>	187.7 - 281.7 3.135 - 4.703	23.5 0.392
CA (AZ III) ABX Pentra Calcium AS CP (A11A01954)	●			mmol/L mg/L mg/dL	<b>2.69</b> <b>107.9</b> <b>10.79</b>	2.15 - 3.23 86.3 - 129.5 8.63 - 12.95	0.27 10.8 1.08
		●		mmol/L mg/L mg/dL	<b>2.69</b> <b>107.9</b> <b>10.79</b>	2.15 - 3.23 86.3 - 129.5 8.63 - 12.95	0.27 10.8 1.08
			●	mmol/L mg/L mg/dL	<b>2.68</b> <b>107.5</b> <b>10.75</b>	2.14 - 3.22 85.9 - 129.1 8.59 - 12.91	0.27 10.8 1.08

\* Confidence range: +/- 2 SD

### For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD		
CREA (CREA3 / CREA4) (Jaffe/Alkaline Picrate) ABX Pentra Creatinine 120 CP (A11A01933)	●			μmol/L mmol/L mg/dL	<b>14151</b> <b>14.15</b> <b>159.9</b>	11321 - 16981 11.31 - 16.99 127.9 - 191.9	1415 1.42 16.0		
		●		μmol/L mmol/L mg/dL	<b>14151</b> <b>14.15</b> <b>159.9</b>	11321 - 16981 11.31 - 16.99 127.9 - 191.9	1415 1.42 16.0		
			●	μmol/L mmol/L mg/dL	<b>12649</b> <b>12.65</b> <b>142.9</b>	10119 - 15179 10.11 - 15.19 114.3 - 171.5	1265 1.27 14.3		
	CREA ENZ. ABX Pentra Enzymatic Creatinine CP (A11A01907)	●			μmol/L mmol/L mg/dL	<b>14966</b> <b>14.97</b> <b>169.1</b>	11972 - 17960 11.97 - 17.97 135.3 - 202.9	1497 1.50 16.9	
			●		μmol/L mmol/L mg/dL	<b>14966</b> <b>14.97</b> <b>169.1</b>	11972 - 17960 11.97 - 17.97 135.3 - 202.9	1497 1.50 16.9	
				●	μmol/L mmol/L mg/dL	<b>13850</b> <b>13.85</b> <b>156.5</b>	11080 - 16620 11.07 - 16.63 125.1 - 187.9	1385 1.39 15.7	
		GluK (HK) ABX Pentra Glucose HK CP (A11A01667)	●			mmol/L g/L mg/dL	<b>16.12</b> <b>2.902</b> <b>290.2</b>	12.90 - 19.34 2.322 - 3.482 232.2 - 348.2	1.61 0.290 29.0
				●		mmol/L g/L mg/dL	<b>16.12</b> <b>2.902</b> <b>290.2</b>	12.90 - 19.34 2.322 - 3.482 232.2 - 348.2	1.61 0.290 29.0
					●	mmol/L g/L mg/dL	<b>15.73</b> <b>2.831</b> <b>283.1</b>	12.59 - 18.87 2.265 - 3.397 226.5 - 339.7	1.57 0.283 28.3
GluP (PAP) ABX Pentra Glucose PAP CP (A11A01668)	●				mmol/L g/L mg/dL	<b>17.51</b> <b>3.152</b> <b>315.2</b>	14.01 - 21.01 2.522 - 3.782 252.2 - 378.2	1.75 0.315 31.5	
			●		mmol/L g/L mg/dL	<b>17.51</b> <b>3.152</b> <b>315.2</b>	14.01 - 21.01 2.522 - 3.782 252.2 - 378.2	1.75 0.315 31.5	
				●	mmol/L g/L mg/dL	<b>16.16</b> <b>2.909</b> <b>290.9</b>	12.92 - 19.40 2.325 - 3.493 232.5 - 349.3	1.62 0.292 29.2	
	PHOS (Ammonium Molybdate UV) ABX Pentra Phosphorus CP (A11A01665)	●			mmol/L mg/L mg/dL	<b>15.25</b> <b>473</b> <b>47.3</b>	12.21 - 18.29 379 - 567 37.9 - 56.7	1.52 47 4.7	
			●		mmol/L mg/L mg/dL	<b>15.25</b> <b>473</b> <b>47.3</b>	12.21 - 18.29 379 - 567 37.9 - 56.7	1.52 47 4.7	
				●	mmol/L mg/L mg/dL	<b>14.82</b> <b>459</b> <b>45.9</b>	11.86 - 17.78 367 - 551 36.7 - 55.1	1.48 46 4.6	

\* Confidence range: +/- 2 SD

### For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD
TPU (Pyrogallol red) ABX Pentra Urinary Proteins CP (A11A01642)	●			g/L	<b>0.637</b>	0.509 - 0.765	0.064
		●		mg/L	<b>637</b>	509 - 765	64
				mg/dL	<b>63.7</b>	50.9 - 76.5	6.4
			●	g/L	<b>0.637</b>	0.509 - 0.765	0.064
		●		mg/L	<b>637</b>	509 - 765	64
				mg/dL	<b>63.7</b>	50.9 - 76.5	6.4
UREA (GLDH) ABX Pentra Urea CP (A11A01641)			●	g/L	<b>0.614</b>	0.492 - 0.736	0.061
				mg/L	<b>614</b>	492 - 736	61
				mg/dL	<b>61.4</b>	49.2 - 73.6	6.1
	●			mmol/L	<b>278</b>	222 - 334	28
		●		mg/L	<b>16680</b>	13320 - 20040	1680
				mg/dL	<b>1668</b>	1332 - 2004	168
BUN (GLDH) ABX Pentra Urea CP (A11A01641)			●	mmol/L	<b>278</b>	222 - 334	28
		●		mg/L	<b>16680</b>	13320 - 20040	1680
				mg/dL	<b>1668</b>	1332 - 2004	168
			●	mmol/L	<b>278</b>	222 - 334	28
		●		mg/L	<b>16680</b>	13320 - 20040	1680
				mg/dL	<b>1668</b>	1332 - 2004	168
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	●			μmol/L	<b>870</b>	696 - 1044	87
				mmol/L	<b>0.87</b>	0.69 - 1.05	0.09
				mg/dL	<b>14.62</b>	11.70 - 17.54	1.46
		●		μmol/L	<b>870</b>	696 - 1044	87
				mmol/L	<b>0.87</b>	0.69 - 1.05	0.09
				mg/dL	<b>14.62</b>	11.70 - 17.54	1.46
			●	μmol/L	<b>890</b>	712 - 1068	89
				mmol/L	<b>0.89</b>	0.71 - 1.07	0.09
				mg/dL	<b>14.95</b>	11.95 - 17.95	1.50

\* Confidence range: +/- 2 SD