



Yumizen G TT for Yumizen G Line



D 4 9 1 0 2 1 4 3 6 3 8 2 0 2 2 3 D



D 5 1 9 4 0 6 9 D

Reference No.:		1300036382
Lot No.:		910214 (rev. 1)
Expiry date:		2023-02-28
MNTT		
Coagulometers		sec
Optical	Yumizen G Line	19,4
Important notice		
The Mean Normal Thrombin Time (MNTT) is for information only. The value depends on the population, race, gender, sampling tube etc. According to the CLSI every laboratory should determine its local MNTT.		