

HORIBA ABX SAS
 Parc Euromédecine
 Rue du Caducée
 B.P. 7290
 34184 Montpellier Cedex 4
 FRANCE
 www.horiba.com




LOT 22002 (rev. 01)



2 2 0 0 2

SAP 1300029568



1300029568

REF : 1300023947



2023/09/30

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400 (P400)	Pentra C400 (PC400)	Pentra C200 (PC200)	Unité	Valeur titrée	Intervalle de confiance	
				Unit	Assigned value	Confidence range	
				Einheit	Deklariertes Wert	Vertrauensintervalle	
				Unidad	Valor asignado	Intervalos de confianza	
				Unità	Valori assegnati	Livelli di fiducia	
				Unidade	Valor determinado	Intervalo de confiança	
				Enhed	Tildelt værdi	Konfidensområde	
				Enhet	Tilldelat värde	Konfidensintervall	
				Μονάδα	Αναγραφόμενη τιμή	Εύρος εμπιστοσύνης	
				Jednostka	Wartość przypisana	Zakres ufności	
				Enhet	Tildelt verd	Konfidensspekter	
C				U	\oplus	Confidence range*	1 SD
AlbT2-U ABX Pentra Micro ALBUMIN 2 CP (1300032563)	●			mg/L mg/dL	63.2 6.32	50.6 - 75.8 5.06 - 7.58	6.3 0.63
		●		mg/L mg/dL	63.2 6.32	50.6 - 75.8 5.06 - 7.58	6.3 0.63
			●	mg/L mg/dL	61.1 6.11	48.9 - 73.3 4.89 - 7.33	6.1 0.61
AMY (EPS-G7) ABX Pentra Amylase CP (A11A01628)	●			U/L μkat/L	248.2 4.145	198.6 - 297.8 3.317 - 4.973	24.8 0.414
		●		U/L μkat/L	248.2 4.145	198.6 - 297.8 3.317 - 4.973	24.8 0.414
			●	U/L μkat/L	236.2 3.945	189.0 - 283.4 3.157 - 4.733	23.6 0.394
CA (AZ III) ABX Pentra Calcium AS CP (A11A01954)	●			mmol/L mg/L mg/dL	2.99 119.9 11.99	2.39 - 3.59 95.9 - 143.9 9.59 - 14.39	0.30 12.0 1.20
		●		mmol/L mg/L mg/dL	2.99 119.9 11.99	2.39 - 3.59 95.9 - 143.9 9.59 - 14.39	0.30 12.0 1.20
			●	mmol/L mg/L mg/dL	2.98 119.5 11.95	2.38 - 3.58 95.5 - 143.5 9.55 - 14.35	0.30 12.0 1.20
	●			mmol/L	2.99	2.39 - 3.59	0.30
		●		mmol/L	2.99	2.39 - 3.59	0.30
			●	mmol/L	2.98	2.38 - 3.58	0.30

* Confidence range: +/- 2 SD

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD	
CREA (CREA3 / CREA4) (Jaffe/Alkaline Picrate) ABX Pentra Creatinine 120 CP (A11A01933)	●			μmol/L mmol/L mg/dL	14241 14.24 160.9	11393 - 17089 11.40 - 17.08 128.7 - 193.1	1424 1.42 16.1	
		●		μmol/L mmol/L mg/dL	14241 14.24 160.9	11393 - 17089 11.40 - 17.08 128.7 - 193.1	1424 1.42 16.1	
			●	μmol/L mmol/L mg/dL	14418 14.42 162.9	11534 - 17302 11.54 - 17.30 130.3 - 195.5	1442 1.44 16.3	
	CREA ENZ. ABX Pentra Enzymatic Creatinine CP (A11A01907)	●			NOT APPLICABLE			
			●		NOT APPLICABLE			
				●	μmol/L mmol/L mg/dL	15276 15.28 172.6	12220 - 18332 12.22 - 18.34 138.0 - 207.2	1528 1.53 17.3
	GluK (HK) ABX Pentra Glucose HK CP (A11A01667)	●			mmol/L g/L mg/dL	15.59 2.806 280.6	12.47 - 18.71 2.244 - 3.368 224.4 - 336.8	1.56 0.281 28.1
			●		mmol/L g/L mg/dL	15.59 2.806 280.6	12.47 - 18.71 2.244 - 3.368 224.4 - 336.8	1.56 0.281 28.1
				●	mmol/L g/L mg/dL	15.54 2.797 279.7	12.44 - 18.64 2.239 - 3.355 223.9 - 335.5	1.55 0.279 27.9
GluP (PAP) ABX Pentra Glucose PAP CP (A11A01668)		●			mmol/L g/L mg/dL	17.75 3.195 319.5	14.21 - 21.29 2.557 - 3.833 255.7 - 383.3	1.77 0.319 31.9
			●		mmol/L g/L mg/dL	17.75 3.195 319.5	14.21 - 21.29 2.557 - 3.833 255.7 - 383.3	1.77 0.319 31.9
				●	mmol/L g/L mg/dL	16.26 2.927 292.7	13.00 - 19.52 2.341 - 3.513 234.1 - 351.3	1.63 0.293 29.3
PHOS (Ammonium Molybdate UV) ABX Pentra Phosphorus CP (A11A01665)	●			mmol/L mg/L mg/dL	15.33 475 47.5	12.27 - 18.39 381 - 569 38.1 - 56.9	1.53 47 4.7	
		●		mmol/L mg/L mg/dL	15.33 475 47.5	12.27 - 18.39 381 - 569 38.1 - 56.9	1.53 47 4.7	
			●	mmol/L mg/L mg/dL	15.25 473 47.3	12.19 - 18.31 379 - 567 37.9 - 56.7	1.53 47 4.7	

* Confidence range: +/- 2 SD

For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD
TPU (Pyrogallol red) ABX Pentra Urinary Proteins CP (A11A01642)	●			g/L mg/L mg/dL	0.604 604 60.4	0.484 - 0.724 484 - 724 48.4 - 72.4	0.060 60 6.0
		●		g/L mg/L mg/dL	0.604 604 60.4	0.484 - 0.724 484 - 724 48.4 - 72.4	0.060 60 6.0
			●	g/L mg/L mg/dL	0.590 590 59.0	0.472 - 0.708 472 - 708 47.2 - 70.8	0.059 59 5.9
UREA (GLDH) ABX Pentra Urea CP (A11A01641)	●			mmol/L mg/L mg/dL	264 15840 1584	212 - 316 12720 - 18960 1272 - 1896	26 1560 156
		●		mmol/L mg/L mg/dL	264 15840 1584	212 - 316 12720 - 18960 1272 - 1896	26 1560 156
			●	mmol/L mg/L mg/dL	246 14760 1476	196 - 296 11760 - 17760 1176 - 1776	25 1500 150
BUN (GLDH) ABX Pentra Urea CP (A11A01641)	●			mmol/L mg/dL	264 741	212 - 316 595 - 887	26 73
		●		mmol/L mg/dL	264 741	212 - 316 595 - 887	26 73
			●	mmol/L mg/dL	246 691	196 - 296 551 - 831	25 70
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	●			μmol/L mmol/L mg/dL	790 0.79 13.27	632 - 948 0.63 - 0.95 10.61 - 15.93	79 0.08 1.33
		●		μmol/L mmol/L mg/dL	790 0.79 13.27	632 - 948 0.63 - 0.95 10.61 - 15.93	79 0.08 1.33
			●	μmol/L mmol/L mg/dL	803 0.80 13.49	643 - 963 0.64 - 0.96 10.81 - 16.17	80 0.08 1.34

* Confidence range: +/- 2 SD