

**HORIBA ABX SAS**  
 Parc Euromédecine  
 Rue du Caducée  
 B.P. 7290  
 34184 Montpellier Cedex 4  
 FRANCE  
 www.horiba.com




**LOT** 22007 (rev. 01)



2 2 0 0 7

**SAP** 1300029568



1300029568

**REF** : 1300023947



2023/09/30

### For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

(FR) Constituant (EN) Constituent (DE) Bestandteil (ES) Componente (IT) Componente (PT) Componente (DA) Bestanddel (SV) Konstituent (EL) συστατικό (PL) Składnik (NO) Bestanddel	ABX Pentra 400 (P400)	Pentra C400 (PC400)	Pentra C200 (PC200)	Unité	Valeur titrée	Intervalle de confiance	
				Unit	Assigned value	Confidence range	
				Einheit	Deklariertes Wert	Vertrauensintervalle	
				Unidad	Valor asignado	Intervalos de confianza	
				Unità	Valori assegnati	Livelli di fiducia	
				Unidade	Valor determinado	Intervalo de confiança	
				Enhed	Tildelt værdi	Konfidensområde	
				Enhet	Tilldelat värde	Konfidensintervall	
				Μονάδα	Αναγραφόμενη τιμή	Εύρος εμπιστοσύνης	
				Jednostka	Wartość przypisana	Zakres ufności	
				Enhet	Tildelt verd	Konfidensspekter	
C				U	$\oplus$	Confidence range*	1 SD
AlbT2-U ABX Pentra Micro ALBUMIN 2 CP (1300032563)	●			mg/L mg/dL	<b>63.2</b> <b>6.32</b>	50.6 - 75.8 5.06 - 7.58	6.3 0.63
		●		mg/L mg/dL	<b>63.2</b> <b>6.32</b>	50.6 - 75.8 5.06 - 7.58	6.3 0.63
			●	mg/L mg/dL	<b>61.1</b> <b>6.11</b>	48.9 - 73.3 4.89 - 7.33	6.1 0.61
AMY (EPS-G7) ABX Pentra Amylase CP (A11A01628)	●			U/L μkat/L	<b>248.2</b> <b>4.145</b>	198.6 - 297.8 3.317 - 4.973	24.8 0.414
		●		U/L μkat/L	<b>248.2</b> <b>4.145</b>	198.6 - 297.8 3.317 - 4.973	24.8 0.414
			●	U/L μkat/L	<b>236.2</b> <b>3.945</b>	189.0 - 283.4 3.157 - 4.733	23.6 0.394
CA (AZ III) ABX Pentra Calcium AS CP (A11A01954)	●			mmol/L mg/L mg/dL	<b>2.99</b> <b>119.9</b> <b>11.99</b>	2.39 - 3.59 95.9 - 143.9 9.59 - 14.39	0.30 12.0 1.20
		●		mmol/L mg/L mg/dL	<b>2.99</b> <b>119.9</b> <b>11.99</b>	2.39 - 3.59 95.9 - 143.9 9.59 - 14.39	0.30 12.0 1.20
			●	mmol/L mg/L mg/dL	<b>2.98</b> <b>119.5</b> <b>11.95</b>	2.38 - 3.58 95.5 - 143.5 9.55 - 14.35	0.30 12.0 1.20
	●			mmol/L mg/L mg/dL	<b>2.99</b> <b>119.9</b> <b>11.99</b>	2.39 - 3.59 95.9 - 143.9 9.59 - 14.39	0.30 12.0 1.20
		●		mmol/L mg/L mg/dL	<b>2.99</b> <b>119.9</b> <b>11.99</b>	2.39 - 3.59 95.9 - 143.9 9.59 - 14.39	0.30 12.0 1.20
			●	mmol/L mg/L mg/dL	<b>2.98</b> <b>119.5</b> <b>11.95</b>	2.38 - 3.58 95.5 - 143.5 9.55 - 14.35	0.30 12.0 1.20

\* Confidence range: +/- 2 SD

### For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD	
CREA (CREA3 / CREA4) (Jaffe/Alkaline Picrate) ABX Pentra Creatinine 120 CP (A11A01933)	●			μmol/L mmol/L mg/dL	<b>14241</b> <b>14.24</b> <b>160.9</b>	11393 - 17089 11.40 - 17.08 128.7 - 193.1	1424 1.42 16.1	
		●		μmol/L mmol/L mg/dL	<b>14241</b> <b>14.24</b> <b>160.9</b>	11393 - 17089 11.40 - 17.08 128.7 - 193.1	1424 1.42 16.1	
			●	μmol/L mmol/L mg/dL	<b>14418</b> <b>14.42</b> <b>162.9</b>	11534 - 17302 11.54 - 17.30 130.3 - 195.5	1442 1.44 16.3	
	CREA ENZ. ABX Pentra Enzymatic Creatinine CP (A11A01907)	●			<b>NOT APPLICABLE</b>			
			●		<b>NOT APPLICABLE</b>			
				●	μmol/L mmol/L mg/dL	<b>15276</b> <b>15.28</b> <b>172.6</b>	12220 - 18332 12.22 - 18.34 138.0 - 207.2	1528 1.53 17.3
	GluK (HK) ABX Pentra Glucose HK CP (A11A01667)	●			mmol/L g/L mg/dL	<b>15.59</b> <b>2.806</b> <b>280.6</b>	12.47 - 18.71 2.244 - 3.368 224.4 - 336.8	1.56 0.281 28.1
			●		mmol/L g/L mg/dL	<b>15.59</b> <b>2.806</b> <b>280.6</b>	12.47 - 18.71 2.244 - 3.368 224.4 - 336.8	1.56 0.281 28.1
				●	mmol/L g/L mg/dL	<b>15.54</b> <b>2.797</b> <b>279.7</b>	12.44 - 18.64 2.239 - 3.355 223.9 - 335.5	1.55 0.279 27.9
GluP (PAP) ABX Pentra Glucose PAP CP (A11A01668)		●			mmol/L g/L mg/dL	<b>17.75</b> <b>3.195</b> <b>319.5</b>	14.21 - 21.29 2.557 - 3.833 255.7 - 383.3	1.77 0.319 31.9
			●		mmol/L g/L mg/dL	<b>17.75</b> <b>3.195</b> <b>319.5</b>	14.21 - 21.29 2.557 - 3.833 255.7 - 383.3	1.77 0.319 31.9
				●	mmol/L g/L mg/dL	<b>16.26</b> <b>2.927</b> <b>292.7</b>	13.00 - 19.52 2.341 - 3.513 234.1 - 351.3	1.63 0.293 29.3
PHOS (Ammonium Molybdate UV) ABX Pentra Phosphorus CP (A11A01665)	●			mmol/L mg/L mg/dL	<b>15.33</b> <b>475</b> <b>47.5</b>	12.27 - 18.39 381 - 569 38.1 - 56.9	1.53 47 4.7	
		●		mmol/L mg/L mg/dL	<b>15.33</b> <b>475</b> <b>47.5</b>	12.27 - 18.39 381 - 569 38.1 - 56.9	1.53 47 4.7	
			●	mmol/L mg/L mg/dL	<b>15.25</b> <b>473</b> <b>47.3</b>	12.19 - 18.31 379 - 567 37.9 - 56.7	1.53 47 4.7	

\* Confidence range: +/- 2 SD

### For instruments ABX Pentra 400 – Pentra C400 – Pentra C200

C	P400	PC400	PC200	U	⊕	Confidence range*	1 SD
TPU (Pyrogallol red) ABX Pentra Urinary Proteins CP (A11A01642)	●			g/L mg/L mg/dL	<b>0.604</b> <b>604</b> <b>60.4</b>	0.484 - 0.724 484 - 724 48.4 - 72.4	0.060 60 6.0
		●		g/L mg/L mg/dL	<b>0.604</b> <b>604</b> <b>60.4</b>	0.484 - 0.724 484 - 724 48.4 - 72.4	0.060 60 6.0
			●	g/L mg/L mg/dL	<b>0.590</b> <b>590</b> <b>59.0</b>	0.472 - 0.708 472 - 708 47.2 - 70.8	0.059 59 5.9
UREA (GLDH) ABX Pentra Urea CP (A11A01641)	●			mmol/L mg/L mg/dL	<b>264</b> <b>15840</b> <b>1584</b>	212 - 316 12720 - 18960 1272 - 1896	26 1560 156
		●		mmol/L mg/L mg/dL	<b>264</b> <b>15840</b> <b>1584</b>	212 - 316 12720 - 18960 1272 - 1896	26 1560 156
			●	mmol/L mg/L mg/dL	<b>246</b> <b>14760</b> <b>1476</b>	196 - 296 11760 - 17760 1176 - 1776	25 1500 150
BUN (GLDH) ABX Pentra Urea CP (A11A01641)	●			mmol/L mg/dL	<b>264</b> <b>741</b>	212 - 316 595 - 887	26 73
		●		mmol/L mg/dL	<b>264</b> <b>741</b>	212 - 316 595 - 887	26 73
			●	mmol/L mg/dL	<b>246</b> <b>691</b>	196 - 296 551 - 831	25 70
UA (Uricase/Peroxidase) ABX Pentra Uric Acid CP (A11A01670)	●			μmol/L mmol/L mg/dL	<b>790</b> <b>0.79</b> <b>13.27</b>	632 - 948 0.63 - 0.95 10.61 - 15.93	79 0.08 1.33
		●		μmol/L mmol/L mg/dL	<b>790</b> <b>0.79</b> <b>13.27</b>	632 - 948 0.63 - 0.95 10.61 - 15.93	79 0.08 1.33
			●	μmol/L mmol/L mg/dL	<b>803</b> <b>0.80</b> <b>13.49</b>	643 - 963 0.64 - 0.96 10.81 - 16.17	80 0.08 1.34

\* Confidence range: +/- 2 SD