

LOT MX 436 L
Rev 1

CONTROL L

 (Exp.)

2022-10-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.0			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.25	2.28			2.20					± 0.17
HB HGB	g/dl	5.8	6.0			5.6					± 0.5
	g/l	58	60			56					± 5
	mmol/l	3.60	3.71			3.47					± 0.31
HT HCT	%	16.2	16.4			16.1					± 2.0
	l/l	0.162	0.164			0.161					± 0.020
VGM MCV	µm ³ ; fl	72	72			73					± 4
TGMH MCH	pg	25.8	26.2			25.4					± 2.0
	fmol	1.60	1.63			1.58					± 0.12
CCMH MCHC	g/dl	35.8	36.4			34.7					± 3.0
	g/l	358	364			347					± 30
	mmol/l	22.23	22.61			21.58					± 1.86
IDR RDW	%	15.5	14.0			15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	75	74			75					± 22
VPM MPV	µm ³ ; fl	8.3	8.1			7.5					± 2.0
LYMPHO	#	1.30	1.31			1.30					± 0.40
	%	64.0	65.7			63.5					± 8.0
MONO	#	0.20	0.20			0.20					± 0.20
	%	8.0	10.0			8.5					± 6.0
GRANULO	#	0.60	0.49			0.60					± 0.40
	%	28.0	24.5			28.0					± 7.0

LOT MX 436 N
 Rev 1

CONTROL N

(Exp.)

2022-10-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.5	7.4			7.3					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.51	4.52			4.49					± 0.26
HB HGB	g/dl	13.5	13.7			13.2					± 0.7
	g/l	135	137			132					± 7
	mmol/l	8.38	8.51			8.22					± 0.43
HT HCT	%	37.4	38.0			38.2					± 2.7
	l/l	0.374	0.380			0.382					± 0.027
VGM MCV	µm ³ ; fl	83	84			85					± 4
TGMH MCH	pg	29.9	30.3			29.5					± 2.0
	fmol	1.86	1.88			1.83					± 0.12
CCMH MCHC	g/dl	36.1	36.1			34.7					± 3.0
	g/l	361	361			347					± 30
	mmol/l	22.40	22.41			21.53					± 1.86
IDR RDW	%	14.5	13.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	263	269			269					± 40
VPM MPV	µm ³ ; fl	8.3	8.0			8.5					± 2.0
LYMPHO	#	2.80	2.82			2.60					± 0.40
	%	37.0	38.2			36.0					± 5.0
MONO	#	0.50	0.53			0.50					± 0.30
	%	6.5	7.1			7.0					± 4.0
GRANULO	#	4.20	4.06			4.20					± 0.60
	%	56.5	54.9			57.0					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

LOT MX 436 H
 Rev 1

CONTROL H

(Exp.)

2022-10-05
 (YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	19.5	19.4			19.3					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.63	5.66			5.58					± 0.32
HB HGB	g/dl	18.2	18.0			17.8					± 0.9
	g/l	182	180			178					± 9
	mmol/l	11.30	11.18			11.07					± 0.56
HT HCT	%	50.1	50.9			50.8					± 3.4
	l/l	0.501	0.509			0.508					± 0.034
VGM MCV	µm ³ ; fl	89	90			91					± 4
TGMH MCH	pg	32.3	31.8			31.9					± 2.0
	fmol	2.01	1.97			1.98					± 0.12
CCMH MCHC	g/dl	36.3	35.3			35.1					± 3.0
	g/l	363	353			351					± 30
	mmol/l	22.56	21.94			21.79					± 1.86
IDR RDW	%	13.5	13.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	490	505			505					± 65
VPM MPV	µm ³ ; fl	8.0	7.6			7.2					± 2.0
LYMPHO	#	3.30	3.29			3.30					± 1.00
	%	17.0	17.0			17.0					± 5.0
MONO	#	1.00	1.26			0.80					± 0.80
	%	5.0	6.5			4.0					± 4.0
GRANULO	#	15.20	14.83			15.20					± 1.40
	%	78.0	76.4			79.0					± 7.0