


**LOT** CX 041  
 Rev 1

**CAL**


 (Exp.) 2011-05-05  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse						TOLERANCES TOLERANCE				
				MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA			PENTRA	PENTRA	PENTRA
				ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60			120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80
GB WBC	$10^9/\text{mm}^3$ ; $10^9/\text{l}$	<b>8.1</b>	<b>8.1</b>	<b>8.1</b>	<b>8.1</b>	<b>8.2</b>		<b>8.3</b>	<b>8.2</b>	<b>8.1</b>	± 0.2			
GR RBC	$10^6/\text{mm}^3$ ; $10^{12}/\text{l}$	<b>4.61</b>	<b>4.62</b>	<b>4.63</b>	<b>4.62</b>	<b>4.61</b>		<b>4.64</b>	<b>4.64</b>	<b>4.62</b>	± 0.06			
HB HGB	g/dl	<b>13.7</b>	<b>13.8</b>	<b>13.8</b>	<b>13.7</b>	<b>13.8</b>		<b>13.8</b>	<b>13.9</b>	<b>13.8</b>	± 0.2			
	g/l	<b>137</b>	<b>138</b>	<b>138</b>	<b>137</b>	<b>138</b>		<b>138</b>	<b>139</b>	<b>138</b>	± 2			
	mmol/l	<b>8.48</b>	<b>8.57</b>	<b>8.57</b>	<b>8.51</b>	<b>8.57</b>		<b>8.57</b>	<b>8.63</b>	<b>8.57</b>	± 0.12			
HT HCT	%	<b>40.2</b>	<b>39.7</b>	<b>40.2</b>	<b>39.7</b>	<b>39.8</b>		<b>38.5</b>	<b>38.4</b>	<b>38.5</b>	± 1.0			
	l/l	<b>0.402</b>	<b>0.397</b>	<b>0.402</b>	<b>0.397</b>	<b>0.398</b>		<b>0.385</b>	<b>0.384</b>	<b>0.385</b>	± 0.010			
PLAQ. PLTS	$10^9/\text{mm}^3$ ; $10^9/\text{l}$	<b>245</b>	<b>245</b>	<b>245</b>	<b>253</b>	<b>250</b>		<b>250</b>	<b>249</b>	<b>250</b>	± 10			
VPM MPV	$\mu\text{m}^3$ ; fl	<b>7.6</b>	<b>7.3</b>	N/A	<b>7.5</b>	<b>7.6</b>		N/A	N/A	N/A	± 0.5			

Ref: FORM-0387 Rev.17 FRONT / RECTO 9930053-A

**LOT** CX 041  
 Rev 1

**CAL**

 (Exp.) 2011-05-05  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l							<b>8.3</b>	<b>8.2</b>	<b>8.1</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l							<b>4.64</b>	<b>4.64</b>	<b>4.62</b>	± 0.06
HB HGB	g/dl							<b>13.8</b>	<b>13.9</b>	<b>13.8</b>	± 0.2
	g/l							<b>138</b>	<b>139</b>	<b>138</b>	± 2
	mmol/l							<b>8.57</b>	<b>8.63</b>	<b>8.57</b>	± 0.12
HT HCT	%							<b>38.5</b>	<b>38.4</b>	<b>38.5</b>	± 1.0
	l/l							<b>0.385</b>	<b>0.384</b>	<b>0.385</b>	± 0.010
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l							<b>250</b>	<b>249</b>	<b>250</b>	± 10
VPM MPV	µm <sup>3</sup> ; fl							N/A	N/A	N/A	± 0.5

Ref: FORM-387 Rev.17 BACK / VERSO 9930053-A