

ABX Minocal



LOT

CX 412
Rev 1

CAL

(Exp.) 2017-08-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.2	9.3	9.1	9.0	9.3					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.52	4.55	4.53	4.50	4.54					± 0.06
HB HGB	g/dl	13.8	13.8	13.6	13.7	13.8					± 0.2
	g/l	138	138	136	137	138					± 2
	mmol/l	8.57	8.57	8.45	8.51	8.57					± 0.12
HT HCT	%	39.1	39.4	39.2	39.2	39.3					± 1.0
	l/l	0.391	0.394	0.392	0.392	0.393					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	235	232	234	238	233					± 10
VPM MPV	µm ³ ; fl	8.2	8.1	N/A	8.2	7.5					± 0.5

Ref: TEMP-0387 Rev.44 FRONT / RECTO 1219930090 -A

ABX Minocal



LOT

CX 412

CAL



(Exp.) 2017-08-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP		MS CRP	XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.5		9.4	9.3	9.5	9.4	9.5	9.3	9.3	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.53		4.50	4.55	4.56	4.54	4.56	4.53	4.55	± 0.06
HB HGB	g/dl	13.7		13.7	13.8	13.7	13.8	13.7	13.8	13.8	± 0.2
	g/l	137		137	138	137	138	137	138	138	± 2
	mmol/l	8.51		8.51	8.57	8.51	8.57	8.51	8.57	8.57	± 0.12
HT HCT	%	39.0		37.0	37.8	38.5	37.5	38.1	37.6	37.8	± 1.0
	l/l	0.390		0.370	0.378	0.385	0.375	0.381	0.376	0.378	± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	237		235	237	237	243	237	240	237	± 10
VPM MPV	µm ³ ; fl	8.3		N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.44 BACK / VERSO 1219930090 -A

ABX Minocal



LOT

CX 412

CAL



(Exp.) 2017-08-05
(YYYY-MM-DD)

Rev **1**

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
		YUMIZEN								
		H1500 H2500								
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.1								± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.57								± 0.06
HB HGB	g/dl	13.6								± 0.2
	g/l	136								± 2
	mmol/l	8.45								± 0.12
HT HCT	%	N/A								N/A
	l/l	N/A								N/A
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	243								± 10
VPM MPV	µm ³ ; fl	N/A								N/A
VGM MCV	µm ³ ; fl	90								± 4

1300032517-A
Ref: TEMP-0387 Rev.44