

ABX Minocal



(Exp.) 2018-05-05

(YYYY - MM - DD)

LOT		CX 421		CAL		(Exp.) 2018-05-05 (YYYY - MM - DD)					
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.0	9.0	8.8	8.9	8.8					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.78	4.78	4.80	4.83	4.80					± 0.06
HB HGB	g/dl	13.8	13.9	13.9	14.0	13.9					± 0.2
	g/l	138	139	139	140	139					± 2
	mmol/l	8.57	8.63	8.63	8.69	8.63					± 0.12
HT HCT	%	39.4	39.4	39.6	40.0	39.8					± 1.0
	l/l	0.394	0.394	0.396	0.400	0.398					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	248	243	236	247	245					± 10
VPM MPV	µm ³ ; fl	8.8	8.6	N/A	9.0	8.0					± 0.5

Ref: TEMP-0387 Rev.45 FRONT / RECTO 1219930090 -A

ABX Minocal


 (Exp.) 2018-05-05
 (YYYY - MM - DD)

LOT		CX 421		CAL		(Exp.) 2018-05-05 (YYYY - MM - DD)							
		Rev 1											
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE		
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA			
		EMI CRP		MS CRP	XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	8.9		8.9	8.7	9.0	8.8	9.0	8.8	8.8	± 0.2		
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.76		4.78	4.80	4.83	4.78	4.83	4.80	4.80	± 0.06		
HB HGB	g/dl	13.7		13.8	13.8	13.8	13.7	13.8	13.8	13.8	± 0.2		
	g/l	137		138	138	138	137	138	138	138	± 2		
	mmol/l	8.51		8.57	8.57	8.57	8.51	8.57	8.57	8.57	± 0.12		
HT HCT	%	39.0		37.5	38.2	38.4	37.5	38.3	38.1	37.9	± 1.0		
	l/l	0.390		0.375	0.382	0.384	0.375	0.383	0.381	0.379	± 0.010		
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	247		238	235	244	248	244	245	245	± 10		
VPM MPV	µm ³ ; fl	8.8		N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5		

 1219930090 -A
 BACK / VERSO
 Ref: TEMP-0387 Rev.45

ABX Minocal


LOT
CX 421
CAL

(Exp.) 2018-05-05

(YYYY - MM - DD)

 Rev **1**

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		YUMIZEN									
		H1500 H2500									
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	8.6									± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.85									± 0.06
HB HGB	g/dl	13.8									± 0.2
	g/l	138									± 2
	mmol/l	8.57									± 0.12
HT HCT	%	N/A									N/A
	l/l	N/A									N/A
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	263									± 10
VPM MPV	µm ³ ; fl	N/A									N/A
VGM MCV	µm ³ ; fl	79.0									± 4.0

1300047683

Ref: TEMP-0387 Rev.45