

ABX Minocal



LOT

CX 400

CAL



(Exp.) 2016-08-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP CRP 200	8P	18P	60 8P 60 18P ADVIA 60					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.1	9.0	9.1	9.0	9.0					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.72	4.65	4.70	4.72	4.66					± 0.06
HB HGB	g/dl	13.7	13.8	13.7	13.8	13.7					± 0.2
	g/l	137	138	137	138	137					± 2
	mmol/l	8.51	8.57	8.51	8.57	8.51					± 0.12
HT HCT	%	39.4	39.1	39.5	39.8	39.1					± 1.0
	l/l	0.394	0.391	0.395	0.398	0.391					± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	245	246	245	246	243					± 10
VPM MPV	µm ³ ; fl	8.7	8.4	N/A	8.7	7.9					± 0.5

Ref: TEMP-0387 Rev.42 FRONT / RECTO 1219930090 -A

ABX Minocal



LOT

CX 400

CAL



(Exp.) 2016-08-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP		MS CRP	XLR	DX NEXUS DF NEXUS	MS60	120 - 120 RETIC DX120 - DF120	60 60 C+ ES60	80 XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.3		9.3	9.2	9.3	9.3	9.3	9.3	9.2	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.68		4.65	4.66	4.70	4.63	4.70	4.68	4.66	± 0.06
HB HGB	g/dl	13.6		13.6	13.7	13.6	13.5	13.6	13.6	13.7	± 0.2
	g/l	136		136	137	136	135	136	136	137	± 2
	mmol/l	8.45		8.45	8.51	8.45	8.38	8.45	8.45	8.51	± 0.12
HT HCT	%	38.8		37.0	37.0	38.0	36.6	38.0	37.3	37.4	± 1.0
	l/l	0.388		0.370	0.370	0.380	0.366	0.380	0.373	0.374	± 0.010
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	246		237	235	250	248	250	250	248	± 10
VPM MPV	µm ³ ; fl	8.6		N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.42 BACK / VERSO 1219930090 -A