

# ABX Minocal



**LOT**

**CX 470**

**CAL**



**(Exp.) 2022-06-05**

Rev 1

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P 60 18P ADVIA 60					
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.3</b>	<b>9.2</b>			<b>9.3</b>					± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.43</b>	<b>4.50</b>			<b>4.47</b>					± 0.06
HB HGB	g/dl	<b>13.8</b>	<b>14.0</b>			<b>13.8</b>					± 0.2
	g/l	<b>138</b>	<b>140</b>			<b>138</b>					± 2
	mmol/l	<b>8.57</b>	<b>8.69</b>			<b>8.57</b>					± 0.12
HT HCT	%	<b>39.2</b>	<b>39.8</b>			<b>39.8</b>					± 1.0
	l/l	<b>0.392</b>	<b>0.398</b>			<b>0.398</b>					± 0.010
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>250</b>	<b>241</b>			<b>246</b>					± 10
VPM MPV	µm <sup>3</sup> ; fl	<b>8.7</b>	<b>8.4</b>			<b>7.9</b>					± 0.5

Ref: TEMP-0387 Rev.47 FRONT / RECTO 1300090308

# ABX Minocal



**LOT**

**CX 470**

**CAL**



**(Exp.) 2022-06-05**  
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>9.3</b>	<b>9.2</b>	<b>9.4</b>	<b>9.3</b>	<b>9.5</b>	<b>9.2</b>	<b>9.5</b>	<b>9.3</b>	<b>9.3</b>	± 0.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.43</b>	<b>4.42</b>	<b>4.35</b>	<b>4.40</b>	<b>4.46</b>	<b>4.46</b>	<b>4.46</b>	<b>4.46</b>	<b>4.40</b>	± 0.06
HB HGB	g/dl	<b>13.6</b>	<b>13.7</b>	<b>13.7</b>	<b>13.8</b>	<b>13.8</b>	<b>13.7</b>	<b>13.8</b>	<b>13.8</b>	<b>13.8</b>	± 0.2
	g/l	<b>136</b>	<b>137</b>	<b>137</b>	<b>138</b>	<b>138</b>	<b>137</b>	<b>138</b>	<b>138</b>	<b>138</b>	± 2
	mmol/l	<b>8.45</b>	<b>8.51</b>	<b>8.51</b>	<b>8.57</b>	<b>8.57</b>	<b>8.51</b>	<b>8.57</b>	<b>8.57</b>	<b>8.57</b>	± 0.12
HT HCT	%	<b>38.8</b>	<b>38.9</b>	<b>37.4</b>	<b>37.8</b>	<b>37.9</b>	<b>37.7</b>	<b>37.9</b>	<b>38.4</b>	<b>37.8</b>	± 1.0
	l/l	<b>0.388</b>	<b>0.389</b>	<b>0.374</b>	<b>0.378</b>	<b>0.379</b>	<b>0.377</b>	<b>0.379</b>	<b>0.384</b>	<b>0.378</b>	± 0.010
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>248</b>	<b>253</b>	<b>241</b>	<b>239</b>	<b>243</b>	<b>245</b>	<b>243</b>	<b>238</b>	<b>239</b>	± 10
VPM MPV	µm <sup>3</sup> ; fl	<b>8.8</b>	<b>9.0</b>	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.47 BACK / VERSO 1300090308

# ABX Minocal



**LOT**

**CX 470**

**CAL**



**(Exp.) 2022-06-05**  
(YYYY-MM-DD)

Rev **1**

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
		YUMIZEN								
		H1500 H2500								
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>8.90</b>								± 0.20
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>4.51</b>								± 0.06
HB HGB	g/dl	<b>13.8</b>								± 0.2
	g/l	<b>138</b>								± 2
	mmol/l	<b>8.57</b>								± 0.12
HT HCT	%	N/A								N/A
	l/l	N/A								N/A
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>250</b>								± 10
VPM MPV	µm <sup>3</sup> ; fl	N/A								N/A
VGM MCV	µm <sup>3</sup> ; fl	<b>87.0</b>								± 4.0

1300060884  
Ref: TEMP-0387 Rev.47