

ABX Minocal



LOT

CX 437

CAL



(Exp.) 2019-09-05

(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|-----------------|--------------|--------------|--------------|-----------------------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | 8P | 18P | 60 8P 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.0 | 8.5 | 8.8 | 8.8 | 8.9 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.46 | 4.48 | 4.45 | 4.45 | 4.46 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.5 | 13.7 | 13.5 | 13.6 | 13.6 | | | | | ± 0.2 |
| | g/l | 135 | 137 | 135 | 136 | 136 | | | | | ± 2 |
| | mmol/l | 8.38 | 8.51 | 8.38 | 8.45 | 8.45 | | | | | ± 0.12 |
| HT HCT | % | 38.3 | 38.8 | 38.5 | 38.7 | 38.6 | | | | | ± 1.0 |
| | l/l | 0.383 | 0.388 | 0.385 | 0.387 | 0.386 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 242 | 242 | 235 | 247 | 244 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 8.0 | 8.0 | N/A | 8.6 | 7.6 | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.46 FRONT / RECTO 1300060880

ABX Minocal



LOT

CX 437

CAL



(Exp.) 2019-09-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------|--------------|--------------|----------------------|--------------|----------------------------------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | 120 - 120 RETIC DX120 - DF120 | 60 60 C+ ES60 | 80 XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.9 | | 9.1 | 8.9 | 9.3 | 8.9 | 9.3 | 8.9 | 8.9 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.45 | | 4.40 | 4.46 | 4.49 | 4.44 | 4.49 | 4.47 | 4.46 | ± 0.06 |
| HB HGB | g/dl | 13.4 | | 13.5 | 13.7 | 13.5 | 13.5 | 13.5 | 13.6 | 13.7 | ± 0.2 |
| | g/l | 134 | | 135 | 137 | 135 | 135 | 135 | 136 | 137 | ± 2 |
| | mmol/l | 8.32 | | 8.38 | 8.51 | 8.38 | 8.38 | 8.38 | 8.45 | 8.51 | ± 0.12 |
| HT HCT | % | 38.3 | | 36.1 | 37.5 | 37.3 | 36.6 | 37.3 | 37.3 | 37.5 | ± 1.0 |
| | l/l | 0.383 | | 0.361 | 0.375 | 0.373 | 0.366 | 0.373 | 0.373 | 0.375 | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 242 | | 233 | 240 | 248 | 243 | 248 | 240 | 240 | ± 10 |
| VPM MPV | µm ³ ; fl | 8.5 | | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.46 BACK / VERSO 1300060880

ABX Minocal



LOT

CX 437

CAL



(Exp.) 2019-09-05

(YYYY-MM-DD)

Rev **1**

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|----------------|--|--|--|--|--|--|--|-------------------------|
| | | YUMIZEN | | | | | | | | |
| | | H1500 H2500 | | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.65 | | | | | | | | ± 0.20 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.51 | | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.5 | | | | | | | | ± 0.2 |
| | g/l | 135 | | | | | | | | ± 2 |
| | mmol/l | 8.38 | | | | | | | | ± 0.12 |
| HT HCT | % | N/A | | | | | | | | N/A |
| | l/l | N/A | | | | | | | | N/A |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 245 | | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | | | | | | | | N/A |
| VGM MCV | µm ³ ; fl | 83.8 | | | | | | | | ± 4.0 |

1300060884

Ref: TEMP-0387 Rev.46