

ABX Minocal



LOT

CX 434

CAL



(Exp.) 2019-06-05

Rev 1

(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|-----------------|--------------|--------------|--------------|-----------------------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | 8P | 18P | 60 8P 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.9 | 8.7 | 8.8 | 8.8 | 8.9 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.46 | 4.48 | 4.49 | 4.47 | 4.50 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.3 | 13.4 | 13.3 | 13.3 | 13.4 | | | | | ± 0.2 |
| | g/l | 133 | 134 | 133 | 133 | 134 | | | | | ± 2 |
| | mmol/l | 8.26 | 8.32 | 8.26 | 8.26 | 8.32 | | | | | ± 0.12 |
| HT HCT | % | 37.8 | 38.1 | 38.2 | 38.0 | 38.3 | | | | | ± 1.0 |
| | l/l | 0.378 | 0.381 | 0.382 | 0.380 | 0.383 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 240 | 240 | 236 | 242 | 240 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 8.7 | 8.6 | N/A | 9.2 | 8.0 | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.46 FRONT / RECTO 1300060880

ABX Minocal



LOT

CX 434
Rev 1

CAL

(Exp.) **2019-06-05**
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------|--------------|--------------|----------------------|--------------|----------------------------------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | 120 - 120 RETIC DX120 - DF120 | 60 60 C+ ES60 | 80 XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.9 | | 9.1 | 8.9 | 9.2 | 8.9 | 9.2 | 8.9 | 8.9 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.46 | | 4.49 | 4.53 | 4.53 | 4.50 | 4.53 | 4.53 | 4.53 | ± 0.06 |
| HB HGB | g/dl | 13.1 | | 13.2 | 13.4 | 13.3 | 13.3 | 13.3 | 13.4 | 13.5 | ± 0.2 |
| | g/l | 131 | | 132 | 134 | 133 | 133 | 133 | 134 | 135 | ± 2 |
| | mmol/l | 8.14 | | 8.20 | 8.32 | 8.26 | 8.26 | 8.26 | 8.32 | 8.38 | ± 0.12 |
| HT HCT | % | 37.5 | | 36.4 | 37.0 | 37.1 | 36.2 | 36.9 | 36.9 | 37.0 | ± 1.0 |
| | l/l | 0.375 | | 0.364 | 0.370 | 0.371 | 0.362 | 0.369 | 0.369 | 0.370 | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 247 | | 237 | 233 | 245 | 242 | 245 | 245 | 244 | ± 10 |
| VPM MPV | µm ³ ; fl | 9.0 | | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.46 BACK / VERSO 1300060880

ABX Minocal



LOT

CX 434

CAL



(Exp.) 2019-06-05

(YYYY-MM-DD)

Rev **1**

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|----------------|--|--|--|--|--|--|--|-------------------------|
| | | YUMIZEN | | | | | | | | |
| | | H1500 H2500 | | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 8.80 | | | | | | | | ± 0.20 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.56 | | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.4 | | | | | | | | ± 0.2 |
| | g/l | 134 | | | | | | | | ± 2 |
| | mmol/l | 8.32 | | | | | | | | ± 0.12 |
| HT HCT | % | N/A | | | | | | | | N/A |
| | l/l | N/A | | | | | | | | N/A |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 245 | | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | | | | | | | | N/A |
| VGM MCV | µm ³ ; fl | 81.8 | | | | | | | | ± 4.0 |

1300060884

Ref: TEMP-0387 Rev.46