

ABX Minocal



LOT

CX 425

CAL



(Exp.) 2018-09-05

(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|---|-----------------|----------------|--------------|--------------|-----------------------------|--------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | PENTRA | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP CRP 200 | 8P | 18P | 60 8P 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.1 | 9.2 | 9.0 | 9.1 | 9.2 | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.63 | 4.67 | 4.62 | 4.62 | 4.63 | | | | | ± 0.06 |
| HB HGB | g/dl | 13.7 | 13.8 | 13.7 | 13.7 | 13.7 | | | | | ± 0.2 |
| | g/l | 137 | 138 | 137 | 137 | 137 | | | | | ± 2 |
| | mmol/l | 8.51 | 8.57 | 8.51 | 8.51 | 8.51 | | | | | ± 0.12 |
| HT HCT | % | 38.9 | 39.4 | 39.1 | 39.0 | 39.1 | | | | | ± 1.0 |
| | l/l | 0.389 | 0.394 | 0.391 | 0.390 | 0.391 | | | | | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 243 | 245 | 232 | 245 | 240 | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | 8.1 | 8.0 | N/A | 8.5 | 7.4 | | | | | ± 0.5 |

Ref: TEMP-0387 Rev.45 FRONT / RECTO 1219930090 -A

ABX Minocal



LOT

CX 425

CAL



(Exp.) 2018-09-05
(YYYY - MM - DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------|--------------|--------------|----------------------|--------------|----------------------------------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | 120 - 120 RETIC DX120 - DF120 | 60 60 C+ ES60 | 80 XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.2 | | 9.1 | 9.1 | 9.3 | 9.2 | 9.3 | 9.2 | 9.1 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.62 | | 4.60 | 4.62 | 4.65 | 4.65 | 4.65 | 4.62 | 4.62 | ± 0.06 |
| HB HGB | g/dl | 13.6 | | 13.6 | 13.7 | 13.7 | 13.7 | 13.7 | 13.6 | 13.7 | ± 0.2 |
| | g/l | 136 | | 136 | 137 | 137 | 137 | 137 | 136 | 137 | ± 2 |
| | mmol/l | 8.45 | | 8.45 | 8.51 | 8.51 | 8.51 | 8.51 | 8.45 | 8.51 | ± 0.12 |
| HT HCT | % | 38.8 | | 37.0 | 37.5 | 37.8 | 37.3 | 37.8 | 37.4 | 37.4 | ± 1.0 |
| | l/l | 0.388 | | 0.370 | 0.375 | 0.378 | 0.373 | 0.378 | 0.374 | 0.374 | ± 0.010 |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 245 | | 240 | 244 | 246 | 254 | 246 | 244 | 244 | ± 10 |
| VPM MPV | µm ³ ; fl | 8.2 | | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.45 BACK / VERSO 1219930090 -A

ABX Minocal



LOT

CX 425

CAL



(Exp.) 2018-09-05
(YYYY-MM-DD)

Rev **1**

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|----------------|--|--|--|--|--|--|--|-------------------------|
| | | YUMIZEN | | | | | | | | |
| | | H1500 H2500 | | | | | | | | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.1 | | | | | | | | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.64 | | | | | | | | ± 0.06 |
| HB HGB | g/dl | 13.7 | | | | | | | | ± 0.2 |
| | g/l | 137 | | | | | | | | ± 2 |
| | mmol/l | 8.51 | | | | | | | | ± 0.12 |
| HT HCT | % | N/A | | | | | | | | N/A |
| | l/l | N/A | | | | | | | | N/A |
| PLAQ. PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 253 | | | | | | | | ± 10 |
| VPM MPV | µm ³ ; fl | N/A | | | | | | | | N/A |
| VGM MCV | µm ³ ; fl | 82.5 | | | | | | | | ± 4.0 |

1300047683
Ref: TEMP-0387 Rev.45